

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN091818\
 Data File : VN051307.D
 Acq On : 18 Sep 2018 19:06
 Operator : MD\SY
 Sample : J4943-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RBR200051

Quant Time: Sep 19 02:02:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091718W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 18 06:04:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	410759	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	630875	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	514371	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	172557	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	228610	51.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.42%	
35) Dibromofluoromethane	7.59	113	235335	48.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.00%	
50) Toluene-d8	10.09	98	812713	44.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.70%	
62) 4-Bromofluorobenzene	12.40	95	217228	37.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.38%	
Target Compounds						
16) Acetone	3.82	43	70378	64.673	ug/l	99
20) Methylene Chloride	4.55	84	23293	4.767	ug/l	97
30) Chloroform	7.37	83	17321	2.133	ug/l	97
43) Isopropyl Acetate	8.17	43	127893	21.054	ug/l #	81
95) Naphthalene	15.13	128	41224	12.654	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN091818\
Data File : VN051307.D
Acq On : 18 Sep 2018 19:06
Operator : MD\SY
Sample : J4943-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RBR200051

Quant Time: Sep 19 02:02:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091718W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 18 06:04:06 2018
Response via : Initial Calibration

