

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN091818\
 Data File : VN051287.D
 Acq On : 18 Sep 2018 9:18
 Operator : MD\SY
 Sample : VN0918WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0918WBL01

Quant Time: Sep 18 14:30:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N082818W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 29 01:38:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	122752	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	645096	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	519234	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	223891	28.58	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	95.27%
60) 4-Bromofluorobenzene	12.40	95	157593	20.41	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	68.03%
63) Toluene-d8	10.09	98	753414	31.12	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	103.73%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN091818\
Data File : VN051287.D
Acq On : 18 Sep 2018 9:18
Operator : MD\SY
Sample : VN0918WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0918WBL01

Quant Time: Sep 18 14:30:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\624N082818W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 29 01:38:44 2018
Response via : Initial Calibration





#1

Bromochloromethane

Concen: 30.000 ug/l

RT: 7.20 min Scan# 1687

Delta R.T. -0.00 min

Lab File: VN051287.D

Acq: 18 Sep 2018 9:18

Instrument :

MSVOA_N

ClientSampled :

VN0918WBL01


Tgt Ion:128 Resp: 122752

Ion Ratio Lower Upper

128 100

49 141.2 0.0 380.3




130 130.6 0.0 325.3



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 49.00 (48.70 to 49.70): VN0

Ion 130.00 (129.70 to 130.70): V



#27

1,2-Dichloroethane-d4

Concen: 28.582 ug/l

RT: 8.03 min Scan# 1945

Delta R.T. -0.00 min

Lab File: VN051287.D


Acq: 18 Sep 2018 9:18

Tgt Ion: 65 Resp: 223891

Ion Ratio Lower Upper



65 100


67 53.8 43.7 65.5



Abundance Ion 65.00 (64.70 to 65.70): VN0

Ion 67.00 (66.70 to 67.70): VN0









#28
 1,4-Difluorobenzene
 Concen: 30.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN051287.D
 Acq: 18 Sep 2018 9:18

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0918WBL01

Tgt Ion:114 Resp: 645096
 Ion Ratio Lower Upper
 114 100
 63 19.0 15.6 23.4
 88 15.2 12.2 18.4




Abundance Ion 114.00 (113.70 to 114.70): V
 Ion 63.00 (62.70 to 63.70): VN
 Ion 88.00 (87.70 to 88.70): VN






#57
 Chlorobenzene-d5
 Concen: 30.000 ug/l
 RT: 11.41 min Scan# 2996
 Delta R.T. -0.00 min
 Lab File: VN051287.D
 Acq: 18 Sep 2018 9:18

Tgt Ion:117 Resp: 519234
 Ion Ratio Lower Upper
 117 100
 82 52.4 42.6 64.0
 119 32.0 25.7 38.5



Abundance Ion 117.00 (116.70 to 117.70): V
 Ion 82.00 (81.70 to 82.70): VN
 Ion 119.00 (118.70 to 119.70): V





#60

4-Bromofluorobenzene

Concen: 20.411 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.00 min

Lab File: VN051287.D

Acq: 18 Sep 2018 9:18

Instrument :

MSVOA_N

ClientSampleId :

VN0918WBL01


Tgt Ion: 95 Resp: 157593

Ion Ratio Lower Upper

95 100

174 95.7 74.6 112.0




176 92.5 71.5 107.3



Abundance Ion 95.00 (94.70 to 95.70): VN051287.D (-3149) (-)

Ion 174.00 (173.70 to 174.70): VN051287.D (-3149) (-)

Ion 176.00 (175.70 to 176.70): VN051287.D (-3149) (-)



#63

Toluene-d8

Concen: 31.115 ug/l

RT: 10.09 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN051287.D


Acq: 18 Sep 2018 9:18

Tgt Ion: 98 Resp: 753414

Ion Ratio Lower Upper

98 100

100 63.5 50.6 75.8



Abundance Ion 98.00 (97.70 to 98.70): VN051287.D (-2431) (-)

Ion 100.00 (99.70 to 100.70): VN051287.D (-2431) (-)

