

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091819\
 Data File : VN058169.D
 Acq On : 18 Sep 2019 17:58
 Operator : JC/SP
 Sample : K4668-11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-01-091719

Quant Time: Sep 19 15:42:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	527403	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	841781	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	734493	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	300968	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	348202	47.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.80%	
35) Dibromofluoromethane	7.58	113	240692	47.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.06%	
50) Toluene-d8	10.09	98	1007652	50.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.08%	
62) 4-Bromofluorobenzene	12.41	95	329067	44.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.86%	
Target Compounds						
16) Acetone	3.79	43	564323	226.325	ug/l	98
20) Methylene Chloride	4.53	84	26122	5.788	ug/l	96
43) Isopropyl Acetate	8.15	43	120762	10.264	ug/l #	87
95) Naphthalene	15.14	128	21671	1.507	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091819\
Data File : VN058169.D
Acq On : 18 Sep 2019 17:58
Operator : JC/SP
Sample : K4668-11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-01-091719

Quant Time: Sep 19 15:42:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

