

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091819\
 Data File : VN058171.D
 Acq On : 18 Sep 2019 18:43
 Operator : JC/SP
 Sample : K4942-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-4

Quant Time: Sep 19 16:40:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	544744	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	906689	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	793938	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	328132	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	376909	49.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.36%	
35) Dibromofluoromethane	7.58	113	258789	46.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.90%	
50) Toluene-d8	10.08	98	1077215	50.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.32%	
62) 4-Bromofluorobenzene	12.40	95	350844	43.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.96%	
Target Compounds						
16) Acetone	3.81	43	12571	4.881	ug/l	92
20) Methylene Chloride	4.53	84	79246	17.431	ug/l	98
43) Isopropyl Acetate	8.15	43	135715	10.709	ug/l #	87
95) Naphthalene	15.13	128	97179	6.200	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091819\
Data File : VN058171.D
Acq On : 18 Sep 2019 18:43
Operator : JC/SP
Sample : K4942-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-4

Quant Time: Sep 19 16:40:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

