

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091819\
 Data File : VN058172.D
 Acq On : 18 Sep 2019 19:05
 Operator : JC/SP
 Sample : K4942-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3

Quant Time: Sep 19 16:43:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	550645	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	886925	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	781631	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	323891	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	373189	48.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.32%	
35) Dibromofluoromethane	7.58	113	253278	46.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.94%	
50) Toluene-d8	10.08	98	1072964	51.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.16%	
62) 4-Bromofluorobenzene	12.41	95	353903	45.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.70%	
Target Compounds						
16) Acetone	3.80	43	15178	5.830	ug/l	95
20) Methylene Chloride	4.53	84	18893	3.941	ug/l	94
43) Isopropyl Acetate	8.15	43	135313	10.915	ug/l #	87
95) Naphthalene	15.14	128	27752	1.794	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091819\
Data File : VN058172.D
Acq On : 18 Sep 2019 19:05
Operator : JC/SP
Sample : K4942-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3

Quant Time: Sep 19 16:43:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

