

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091819\
 Data File : VN058175.D
 Acq On : 18 Sep 2019 20:12
 Operator : JC/SP
 Sample : K4931-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 290-371

Quant Time: Sep 19 16:50:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	534479	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	880302	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	764441	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	319988	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	361066	48.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.00%	
35) Dibromofluoromethane	7.58	113	247961	46.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.66%	
50) Toluene-d8	10.09	98	1045734	50.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.30%	
62) 4-Bromofluorobenzene	12.40	95	339030	43.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.54%	
Target Compounds						
16) Acetone	3.80	43	32796	12.979	ug/l	96
20) Methylene Chloride	4.53	84	30781	6.766	ug/l	99
43) Isopropyl Acetate	8.16	43	124957	10.156	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091819\
Data File : VN058175.D
Acq On : 18 Sep 2019 20:12
Operator : JC/SP
Sample : K4931-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
290-371

Quant Time: Sep 19 16:50:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

