

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091819\
 Data File : VN058177.D
 Acq On : 18 Sep 2019 20:56
 Operator : JC/SP
 Sample : K4944-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DRUM-COMP

Quant Time: Sep 19 16:57:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	534187	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	873120	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	764644	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	336347	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	365279	49.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.20%	
35) Dibromofluoromethane	7.58	113	253989	47.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.70%	
50) Toluene-d8	10.09	98	1062318	51.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.74%	
62) 4-Bromofluorobenzene	12.40	95	348289	45.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.68%	
Target Compounds						
16) Acetone	3.80	43	16355	6.476	ug/l	96
20) Methylene Chloride	4.53	84	2132389	484.190	ug/l	98
68) m/p-Xylenes	11.62	106	20812	2.398	ug/l	97
69) o-Xylene	11.95	106	58262	6.827	ug/l	99
80) 1,3,5-Trimethylbenzene	12.74	105	205332	11.534	ug/l	98
84) 1,2,4-Trimethylbenzene	13.05	105	344904	19.468	ug/l	99
86) p-Isopropyltoluene	13.29	119	26776	1.451	ug/l	97
95) Naphthalene	15.13	128	20372	1.268	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091819\
Data File : VN058177.D
Acq On : 18 Sep 2019 20:56
Operator : JC/SP
Sample : K4944-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DRUM-COMP

Quant Time: Sep 19 16:57:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

