


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN091819\
Data File : VN058159.D
Acq On : 18 Sep 2019 13:32
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Title : SW846 8260
Last Update : Thu Sep 19 09:27:53 2019



AutoFind: Scans 3305, 3306, 3307; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.7	31221	PASS
75	95	30	60	55.0	75613	PASS
95	95	100	100	100.0	137448	PASS
96	95	5	9	6.5	8925	PASS
173	174	0.00	2	1.1	1184	PASS
174	95	50	100	75.2	103354	PASS
175	174	5	9	7.3	7534	PASS
176	174	95	101	95.0	98184	PASS
177	176	5	9	7.0	6895	PASS