

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091820\
 Data File : VN063449.D
 Acq On : 18 Sep 2020 11:36
 Operator : JC/MD
 Sample : VN0918WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0918WBL01

Quant Time: Sep 21 01:46:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	254906	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	471970	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	466970	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	201335	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	205794	47.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.86%	
35) Dibromofluoromethane	7.55	113	150573	46.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.16%	
50) Toluene-d8	10.06	98	607932	49.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.96%	
62) 4-Bromofluorobenzene	12.38	95	242976	53.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.90%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091820\
Data File : VN063449.D
Acq On : 18 Sep 2020 11:36
Operator : JC/MD
Sample : VN0918WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0918WBL01

Quant Time: Sep 21 01:46:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

