

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091820\
 Data File : VN063456.D
 Acq On : 18 Sep 2020 14:38
 Operator : JC/MD
 Sample : PB131701ZHE#11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131701ZHE#11

Quant Time: Sep 21 02:00:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	247881	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	459636	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	466446	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	202689	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	199673	47.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.64%	
35) Dibromofluoromethane	7.55	113	147450	46.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.68%	
50) Toluene-d8	10.06	98	591564	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
62) 4-Bromofluorobenzene	12.38	95	240450	54.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.64%	
Target Compounds						
16) Acetone	3.78	43	32347	23.853	ug/l	98
18) Methyl Acetate	4.28	43	5215	1.515	ug/l #	75
20) Methylene Chloride	4.51	84	11458	3.335	ug/l	94
43) Isopropyl Acetate	8.13	43	100686	11.716	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091820\
Data File : VN063456.D
Acq On : 18 Sep 2020 14:38
Operator : JC/MD
Sample : PB131701ZHE#11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131701ZHE#11

Quant Time: Sep 21 02:00:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

