

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091820\
 Data File : VN063458.D
 Acq On : 18 Sep 2020 15:26
 Operator : JC/MD
 Sample : PB131701ZHE#13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131701ZHE#13

Manual Integrations
 APPROVED

MMDadoda
 9/21/2020 11:49:44 AM

Quant Time: Sep 21 02:06:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	252913	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	464579	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	471044	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	204697	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	204646	47.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.06%	
35) Dibromofluoromethane	7.55	113	146206	45.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.92%	
50) Toluene-d8	10.06	98	606387	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
62) 4-Bromofluorobenzene	12.38	95	244078	55.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.12%	
Target Compounds						
16) Acetone	3.78	43	30691	22.181	ug/l	98
18) Methyl Acetate	4.28	43	3819m	1.087	ug/l	
20) Methylene Chloride	4.51	84	11681	3.332	ug/l	96
43) Isopropyl Acetate	8.13	43	96772	11.141	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091820\
Data File : VN063458.D
Acq On : 18 Sep 2020 15:26
Operator : JC/MD
Sample : PB131701ZHE#13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131701ZHE#13

Manual Integrations
APPROVED

MMDadoda
9/21/2020 11:49:44 AM

Quant Time: Sep 21 02:06:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

