

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091820\
 Data File : VN063460.D
 Acq On : 18 Sep 2020 16:13
 Operator : JC/MD
 Sample : PB131701ZHE#15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131701ZHE#15

Manual Integrations
 APPROVED

MMDadoda
 9/21/2020 11:49:46 AM

Quant Time: Sep 21 02:10:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	243666	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	457822	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	457225	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	202290	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	201534	48.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.18%	
35) Dibromofluoromethane	7.55	113	146526	46.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.46%	
50) Toluene-d8	10.06	98	590255	49.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.06%	
62) 4-Bromofluorobenzene	12.38	95	239471	54.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.62%	
Target Compounds						
16) Acetone	3.78	43	39065	29.305	ug/l	97
18) Methyl Acetate	4.28	43	5131m	1.516	ug/l	
20) Methylene Chloride	4.51	84	21430	6.345	ug/l	96
43) Isopropyl Acetate	8.13	43	89138	10.414	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091820\
Data File : VN063460.D
Acq On : 18 Sep 2020 16:13
Operator : JC/MD
Sample : PB131701ZHE#15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131701ZHE#15

Manual Integrations
APPROVED

MMDadoda
9/21/2020 11:49:46 AM

Quant Time: Sep 21 02:10:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

