

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091820\
 Data File : VN063461.D
 Acq On : 18 Sep 2020 16:37
 Operator : JC/MD
 Sample : PB131701ZHE#16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131701ZHE#16

Manual Integrations
 APPROVED

MMDadoda
 9/21/2020 11:49:48 AM

Quant Time: Sep 21 02:12:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	234742	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	446218	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	453278	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	191789	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	194220	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
35) Dibromofluoromethane	7.55	113	141477	45.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.60%	
50) Toluene-d8	10.06	98	581637	50.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.14%	
62) 4-Bromofluorobenzene	12.38	95	228347	53.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.26%	
Target Compounds						
16) Acetone	3.78	43	53734	41.842	ug/l	96
18) Methyl Acetate	4.29	43	5343m	1.639	ug/l	
20) Methylene Chloride	4.51	84	32516	9.993	ug/l	97
43) Isopropyl Acetate	8.13	43	82756	9.920	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091820\
Data File : VN063461.D
Acq On : 18 Sep 2020 16:37
Operator : JC/MD
Sample : PB131701ZHE#16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131701ZHE#16

Manual Integrations
APPROVED

MMDadoda
9/21/2020 11:49:48 AM

Quant Time: Sep 21 02:12:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

