

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091820\
 Data File : VN063464.D
 Acq On : 18 Sep 2020 17:49
 Operator : JC/MD
 Sample : L4059-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAV-B33-091720-0-1

Quant Time: Sep 21 02:20:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	250681	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	477898	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	488449	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	213246	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	211563	49.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.16%	
35) Dibromofluoromethane	7.55	113	153372	46.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.72%	
50) Toluene-d8	10.06	98	623609	50.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.26%	
62) 4-Bromofluorobenzene	12.38	95	254872	55.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.78%	
Target Compounds						
16) Acetone	3.78	43	31860	23.231	ug/l	97
18) Methyl Acetate	4.29	43	4161	1.195	ug/l #	75
20) Methylene Chloride	4.51	84	14830	4.268	ug/l	95
43) Isopropyl Acetate	8.13	43	90835	10.166	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091820\
Data File : VN063464.D
Acq On : 18 Sep 2020 17:49
Operator : JC/MD
Sample : L4059-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAV-B33-091720-0-1

Quant Time: Sep 21 02:20:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

