

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091820\
 Data File : VN063466.D
 Acq On : 18 Sep 2020 18:37
 Operator : JC/MD
 Sample : L3870-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-02-091720

Manual Integrations
 APPROVED

MMDadoda
 9/21/2020 11:49:52 AM

Quant Time: Sep 21 02:25:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	232617	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	434506	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	441906	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	191254	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	190182	48.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.06%	
35) Dibromofluoromethane	7.55	113	140624	46.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.50%	
50) Toluene-d8	10.06	98	562186	49.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.40%	
62) 4-Bromofluorobenzene	12.38	95	225266	54.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.66%	
Target Compounds						
16) Acetone	3.78	43	46515	36.551	ug/l	98
18) Methyl Acetate	4.29	43	5323m	1.648	ug/l	
20) Methylene Chloride	4.51	84	13422	4.163	ug/l	97
43) Isopropyl Acetate	8.13	43	76730	9.445	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091820\
Data File : VN063466.D
Acq On : 18 Sep 2020 18:37
Operator : JC/MD
Sample : L3870-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-02-091720

Manual Integrations
APPROVED

MMDadoda
9/21/2020 11:49:52 AM

Quant Time: Sep 21 02:25:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

