

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091820\
 Data File : VN063469.D
 Acq On : 18 Sep 2020 19:48
 Operator : JC/MD
 Sample : L4075-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 KTS-WC-S-IB-ST

Quant Time: Sep 21 02:34:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	234053	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	447931	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	452185	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	195752	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	197859	49.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.32%	
35) Dibromofluoromethane	7.55	113	140791	45.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.80%	
50) Toluene-d8	10.06	98	574453	49.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.54%	
62) 4-Bromofluorobenzene	12.38	95	229787	53.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.52%	
Target Compounds						
16) Acetone	3.78	43	60606	47.332	ug/l	98
18) Methyl Acetate	4.29	43	4246	1.306	ug/l #	80
20) Methylene Chloride	4.51	84	9833	3.031	ug/l	86
43) Isopropyl Acetate	8.13	43	74402	8.884	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091820\
Data File : VN063469.D
Acq On : 18 Sep 2020 19:48
Operator : JC/MD
Sample : L4075-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
KTS-WC-S-IB-ST

Quant Time: Sep 21 02:34:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

