

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091820\
 Data File : VN063470.D
 Acq On : 18 Sep 2020 20:12
 Operator : JC/MD
 Sample : PB131708TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131708TB

Quant Time: Sep 21 02:35:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	222691	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	422715	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	432411	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	184566	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	187859	49.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.12%	
35) Dibromofluoromethane	7.55	113	136238	46.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.12%	
50) Toluene-d8	10.06	98	549032	49.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.80%	
62) 4-Bromofluorobenzene	12.38	95	219808	54.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.98%	
Target Compounds						
16) Acetone	3.78	43	30737	25.230	ug/l	98
18) Methyl Acetate	4.29	43	4439	1.435	ug/l	98
20) Methylene Chloride	4.51	84	10848	3.514	ug/l	89
43) Isopropyl Acetate	8.13	43	92250	11.672	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091820\
Data File : VN063470.D
Acq On : 18 Sep 2020 20:12
Operator : JC/MD
Sample : PB131708TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131708TB

Quant Time: Sep 21 02:35:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

