

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091823\
 Data File : VN079286.D
 Acq On : 18 Sep 2023 12:57
 Operator : JC\MD
 Sample : VN0918MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0918MBL01

Quant Time: Sep 19 01:22:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	146434	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	276687	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	261283	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	70520	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	126551	64.021	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 128.040%#	
35) Dibromofluoromethane	8.171	113	103354	56.156	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 112.320%	
50) Toluene-d8	10.571	98	350357	51.844	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.680%	
62) 4-Bromofluorobenzene	12.859	95	101271	46.979	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.960%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091823\
Data File : VN079286.D
Acq On : 18 Sep 2023 12:57
Operator : JC\MD
Sample : VN0918MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0918MBL01

Quant Time: Sep 19 01:22:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

