

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091823\
 Data File : VN079291.D
 Acq On : 18 Sep 2023 17:37
 Operator : JC\MD
 Sample : 04414-02RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-01-091223RE

Quant Time: Sep 19 01:26:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	151877	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	279992	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	238251	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	62059	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	110982	54.132	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.260%
35) Dibromofluoromethane	8.171	113	84000	45.102	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.200%
50) Toluene-d8	10.571	98	274509	40.141	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	80.280%#
62) 4-Bromofluorobenzene	12.853	95	77786	35.659	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	71.320%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091823\
Data File : VN079291.D
Acq On : 18 Sep 2023 17:37
Operator : JC\MD
Sample : 04414-02RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-01-091223RE

Quant Time: Sep 19 01:26:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

