

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091823\
 Data File : VN079292.D
 Acq On : 18 Sep 2023 18:01
 Operator : JC\MD
 Sample : 04436-21RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB-9-13RE

Quant Time: Sep 19 01:26:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	136768	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	256970	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	220787	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	56417	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	108095	58.549	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.100%
35) Dibromofluoromethane	8.171	113	82365	48.186	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.380%
50) Toluene-d8	10.571	98	275352	43.871	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	87.740%
62) 4-Bromofluorobenzene	12.853	95	76155	38.039	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	76.080%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091823\
Data File : VN079292.D
Acq On : 18 Sep 2023 18:01
Operator : JC\MD
Sample : 04436-21RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB-9-13RE

Quant Time: Sep 19 01:26:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

