

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091823\
 Data File : VN079293.D
 Acq On : 18 Sep 2023 18:25
 Operator : JC\MD
 Sample : 04436-22RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-9-13RE

Quant Time: Sep 19 01:27:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :John Carlone 09/19/2023
 Supervised By :Semsettin Yesilyurt 09/19/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	133430	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	254700	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	215471	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	55108	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	104587	58.066	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.140%
35) Dibromofluoromethane	8.171	113	82950	48.961	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.920%
50) Toluene-d8	10.571	98	274088	44.059	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.120%
62) 4-Bromofluorobenzene	12.853	95	76880	38.743	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	77.480%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091823\
Data File : VN079293.D
Acq On : 18 Sep 2023 18:25
Operator : JC\MD
Sample : 04436-22RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-9-13RE

Quant Time: Sep 19 01:27:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 09/19/2023
Supervised By :Semsettin Yesilyurt 09/19/2023

