

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091824\
 Data File : VN083987.D
 Acq On : 18 Sep 2024 17:03
 Operator : JC\MD
 Sample : PB163447ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB163447ZHE#04

Manual Integrations
 APPROVED

Reviewed By : John Carlone 09/19/2024
 Supervised By : Mahesh Dadoda 09/19/2024

Quant Time: Sep 19 01:36:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	178438	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	310028	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	262435	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	109083	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	135885	50.504	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 101.000%			
35) Dibromofluoromethane	8.171	113	95886	48.311	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 96.620%			
50) Toluene-d8	10.565	98	364650	53.145	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 106.280%			
62) 4-Bromofluorobenzene	12.847	95	128041	45.231	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 90.460%			
Target Compounds					Qvalue	
16) Acetone	4.436	43	21830	20.239	ug/l	94
20) Methylene Chloride	5.283	84	6559	2.972	ug/l	95
43) Isopropyl Acetate	8.688	43	289866m	58.990	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091824\
Data File : VN083987.D
Acq On : 18 Sep 2024 17:03
Operator : JC\MD
Sample : PB163447ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB163447ZHE#04

Manual Integrations
APPROVED

Reviewed By : John Carlone 09/19/2024
Supervised By : Mahesh Dadoda 09/19/2024

Quant Time: Sep 19 01:36:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

