

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091824\
 Data File : VN083989.D
 Acq On : 18 Sep 2024 17:51
 Operator : JC\MD
 Sample : PB163447ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB163447ZHE#06

Manual Integrations
 APPROVED

Reviewed By : John Carlone 09/19/2024
 Supervised By : Mahesh Dadoda 09/19/2024

Quant Time: Sep 19 01:38:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	168573	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	287714	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	248320	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	103718	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	129237	50.844	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 101.680%			
35) Dibromofluoromethane	8.165	113	90035	48.881	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.760%			
50) Toluene-d8	10.565	98	347531	54.578	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 109.160%			
62) 4-Bromofluorobenzene	12.847	95	123425	46.981	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 93.960%			
Target Compounds						
16) Acetone	4.442	43	21718	21.314	ug/l	100
20) Methylene Chloride	5.277	84	6747	3.236	ug/l #	81
43) Isopropyl Acetate	8.688	43	276460m	60.625	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091824\
Data File : VN083989.D
Acq On : 18 Sep 2024 17:51
Operator : JC\MD
Sample : PB163447ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB163447ZHE#06

Manual Integrations
APPROVED

Reviewed By :John Carlone 09/19/2024
Supervised By :Mahesh Dadoda 09/19/2024

Quant Time: Sep 19 01:38:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

