

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091824\
 Data File : VN083991.D
 Acq On : 18 Sep 2024 18:39
 Operator : JC\MD
 Sample : PB163447ZHE#08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB163447ZHE#08

Manual Integrations
 APPROVED

Reviewed By :John
 Carlone

09/19/2024
 Supervised By :Mahesh
 Dadoda

Quant Time: Sep 19 01:39:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	D
-----							05
Internal Standards							
1) Pentafluorobenzene	8.224	168	169652	50.000	ug/l	0.00	
34) 1,4-Difluorobenzene	9.100	114	293496	50.000	ug/l	0.00	
63) Chlorobenzene-d5	11.865	117	252091	50.000	ug/l	0.00	
72) 1,4-Dichlorobenzene-d4	13.794	152	106530	50.000	ug/l	0.00	
System Monitoring Compounds							
33) 1,2-Dichloroethane-d4	8.577	65	131193	51.285	ug/l	0.00	
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.580%	
35) Dibromofluoromethane	8.165	113	91685	48.797	ug/l	0.00	
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.600%	
50) Toluene-d8	10.565	98	354770	54.617	ug/l	0.00	
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.240%	
62) 4-Bromofluorobenzene	12.847	95	123669	46.147	ug/l	0.00	
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.300%	
Target Compounds							
16) Acetone	4.430	43	44300	43.199	ug/l	91	
20) Methylene Chloride	5.283	84	3560	1.696	ug/l #	78	
43) Isopropyl Acetate	8.688	43	236683m	50.880	ug/l		

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091824\
Data File : VN083991.D
Acq On : 18 Sep 2024 18:39
Operator : JC\MD
Sample : PB163447ZHE#08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB163447ZHE#08

Manual Integrations
APPROVED

Reviewed By :John
Carlone

09/19/2024
Supervised By :Mahesh
Dadoda

09/19/2024

Quant Time: Sep 19 01:39:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

