

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091824\
 Data File : VN083996.D
 Acq On : 18 Sep 2024 20:39
 Operator : JC\MD
 Sample : PB163447ZHE#13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB163447ZHE#13

Manual Integrations
 APPROVED

Reviewed By : John Carlone 09/19/2024
 Supervised By : Mahesh Dadoda 09/19/2024

Quant Time: Sep 19 01:41:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	165216	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	294509	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	253848	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	102465	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	127590	51.216	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 102.440%			
35) Dibromofluoromethane	8.165	113	90807	48.163	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 96.320%			
50) Toluene-d8	10.565	98	350905	53.837	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 107.680%			
62) 4-Bromofluorobenzene	12.847	95	121646	45.236	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 90.480%			
Target Compounds					Qvalue	
16) Acetone	4.430	43	43089	43.146	ug/l	96
20) Methylene Chloride	5.277	84	4043	1.978	ug/l	88
43) Isopropyl Acetate	8.688	43	246406m	52.788	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091824\
Data File : VN083996.D
Acq On : 18 Sep 2024 20:39
Operator : JC\MD
Sample : PB163447ZHE#13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB163447ZHE#13

Manual Integrations
APPROVED

Reviewed By :John Carlone 09/19/2024
Supervised By :Mahesh Dadoda 09/19/2024

Quant Time: Sep 19 01:41:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

