

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091824\
 Data File : VN084007.D
 Acq On : 19 Sep 2024 03:22
 Operator : JC\MD
 Sample : VN0918WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0918WBL02

Quant Time: Sep 19 05:46:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	169005	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	290063	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	233833	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	96839	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	128159	50.291	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.580%	
35) Dibromofluoromethane	8.171	113	92672	49.906	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.820%	
50) Toluene-d8	10.565	98	336990	52.494	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.980%	
62) 4-Bromofluorobenzene	12.847	95	120864	45.634	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 91.260%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091824\
Data File : VN084007.D
Acq On : 19 Sep 2024 03:22
Operator : JC\MD
Sample : VN0918WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0918WBL02

Quant Time: Sep 19 05:46:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

