

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091824\
 Data File : VN084015.D
 Acq On : 19 Sep 2024 06:33
 Operator : JC\MD
 Sample : P4087-26
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NB-614-COMP-05

Manual Integrations
 APPROVED

Reviewed By : John Carlone 09/19/2024
 Supervised By : Mahesh Dadoda 09/19/2024

Quant Time: Sep 19 07:03:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	169096	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	298233	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	257179	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	103882	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	132712	52.049	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.100%			
35) Dibromofluoromethane	8.165	113	94270	49.375	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 98.760%			
50) Toluene-d8	10.565	98	355555	53.869	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 107.740%			
62) 4-Bromofluorobenzene	12.847	95	129453	47.538	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 95.080%			
Target Compounds						
16) Acetone	4.430	43	31578	30.894	ug/l	94
20) Methylene Chloride	5.283	84	4450	2.128	ug/l	94
25) 2-Butanone	7.488	43	15620	11.627	ug/l	98
43) Isopropyl Acetate	8.682	43	229572m	48.567	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091824\
Data File : VN084015.D
Acq On : 19 Sep 2024 06:33
Operator : JC\MD
Sample : P4087-26
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NB-614-COMP-05

Manual Integrations
APPROVED

Reviewed By : John Carlone 09/19/2024
Supervised By : Mahesh Dadoda 09/19/2024

Quant Time: Sep 19 07:03:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

