

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091824\
 Data File : VN083975.D
 Acq On : 18 Sep 2024 11:55
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 19 01:29:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.581	0.531	8.6	84	0.00
3 P	Chloromethane	0.707	0.522	26.2#	80	0.00
4 C	Vinyl Chloride	0.636	0.554	12.9#	81	0.00
5 T	Bromomethane	0.365	0.364	0.3	97	0.04
6 T	Chloroethane	0.446	0.375	15.9	83	0.03
7 T	Trichlorofluoromethane	1.080	0.992	8.1	89	0.02
8 T	Diethyl Ether	0.362	0.358	1.1	93	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.573	0.534	6.8	87	0.02
10 T	Methyl Iodide	0.728	0.694	4.7	87	0.01
11 T	Tert butyl alcohol	0.101	0.123	-21.8	116	-0.01
12 CM	1,1-Dichloroethene	0.541	0.515	4.8#	88	0.01
13 T	Acrolein	0.100	0.092	8.0	85	0.00
14 T	Allyl chloride	0.964	0.871	9.6	83	0.00
15 T	Acrylonitrile	0.275	0.279	-1.5	97	0.00
16 T	Acetone	0.302	0.291	3.6	93	0.00
17 T	Carbon Disulfide	1.513	1.261	16.7	79	0.01
18 T	Methyl Acetate	0.735	0.711	3.3	102	0.00
19 T	Methyl tert-butyl Ether	1.957	1.988	-1.6	95	0.00
20 T	Methylene Chloride	0.618	0.579	6.3	91	0.00
21 T	trans-1,2-Dichloroethene	0.580	0.530	8.6	87	0.00
22 T	Diisopropyl ether	1.984	2.013	-1.5	94	0.00
23 T	Vinyl Acetate	1.453	1.497	-3.0	92	0.00
24 P	1,1-Dichloroethane	1.143	1.107	3.1	92	0.00
25 T	2-Butanone	0.397	0.407	-2.5	99	0.00
26 T	2,2-Dichloropropane	1.052	1.022	2.9	89	0.00
27 T	cis-1,2-Dichloroethene	0.688	0.682	0.9	91	0.00
28 T	Bromochloromethane	0.481	0.490	-1.9	89	0.00
29 T	Tetrahydrofuran	0.235	0.253	-7.7	100	0.00
30 C	Chloroform	1.213	1.181	2.6#	92	0.00
31 T	Cyclohexane	1.090	0.907	16.8	85	0.00
32 T	1,1,1-Trichloroethane	1.116	1.095	1.9	92	0.00
33 S	1,2-Dichloroethane-d4	0.754	0.759	-0.7	85	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.320	0.320	0.0	89	0.00
36 T	1,1-Dichloropropene	0.482	0.456	5.4	91	0.00
37 T	Ethyl Acetate	0.437	0.481	-10.1	99	0.00
38 T	Carbon Tetrachloride	0.547	0.535	2.2	90	0.00
39 T	Methylcyclohexane	0.548	0.506	7.7	83	0.00
40 TM	Benzene	1.450	1.388	4.3	91	0.00
41 T	Methacrylonitrile	0.259	0.269	-3.9	96	0.00
42 TM	1,2-Dichloroethane	0.542	0.547	-0.9	94	0.00
43 T	Isopropyl Acetate	0.792	0.850	-7.3	100	0.00
44 TM	Trichloroethene	0.336	0.328	2.4	91	0.00
45 C	1,2-Dichloropropane	0.350	0.348	0.6#	94	0.00
46 T	Dibromomethane	0.245	0.249	-1.6	92	0.00
47 T	Bromodichloromethane	0.531	0.546	-2.8	94	0.00
48 T	Methyl methacrylate	0.376	0.414	-10.1	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091824\
 Data File : VN083975.D
 Acq On : 18 Sep 2024 11:55
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 19 01:29:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.007	-16.7	105	0.00
50 S	Toluene-d8	1.107	1.141	-3.1	84	0.00
51 T	4-Methyl-2-Pentanone	0.431	0.484	-12.3	103	0.00
52 CM	Toluene	0.887	0.874	1.5#	91	0.00
53 T	t-1,3-Dichloropropene	0.539	0.565	-4.8	95	0.00
54 T	cis-1,3-Dichloropropene	0.566	0.585	-3.4	94	0.00
55 T	1,1,2-Trichloroethane	0.331	0.332	-0.3	95	0.00
56 T	Ethyl methacrylate	0.509	0.576	-13.2	97	0.00
57 T	1,3-Dichloropropane	0.569	0.583	-2.5	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.184	0.242	-31.5#	132	0.00
59 T	2-Hexanone	0.319	0.357	-11.9	102	0.00
60 T	Dibromochloromethane	0.371	0.401	-8.1	96	0.00
61 T	1,2-Dibromoethane	0.330	0.339	-2.7	96	0.00
62 S	4-Bromofluorobenzene	0.457	0.452	1.1	87	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.333	0.320	3.9	94	0.00
65 PM	Chlorobenzene	1.111	1.082	2.6	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.385	0.393	-2.1	94	0.00
67 C	Ethyl Benzene	1.974	2.034	-3.0#	94	0.00
68 T	m/p-Xylenes	0.733	0.749	-2.2	93	0.00
69 T	o-Xylene	0.710	0.726	-2.3	92	0.00
70 T	Styrene	1.164	1.248	-7.2	94	0.00
71 P	Bromoform	0.260	0.291	-11.9	99	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	3.868	3.924	-1.4	92	0.00
74 T	N-amyl acetate	1.650	1.802	-9.2	99	0.00
75 P	1,1,2,2-Tetrachloroethane	1.137	1.125	1.1	98	0.00
76 T	1,2,3-Trichloropropane	1.009	0.982	2.7	99	0.00
77 T	Bromobenzene	0.931	0.886	4.8	94	0.00
78 T	n-propylbenzene	4.501	4.617	-2.6	92	0.00
79 T	2-Chlorotoluene	2.870	2.875	-0.2	95	0.00
80 T	1,3,5-Trimethylbenzene	3.261	3.349	-2.7	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.385	0.401	-4.2	93	0.00
82 T	4-Chlorotoluene	2.880	2.897	-0.6	94	0.00
83 T	tert-Butylbenzene	2.853	2.932	-2.8	93	0.00
84 T	1,2,4-Trimethylbenzene	3.276	3.366	-2.7	92	0.00
85 T	sec-Butylbenzene	3.784	3.943	-4.2	92	0.00
86 T	p-Isopropyltoluene	3.172	3.298	-4.0	91	0.00
87 T	1,3-Dichlorobenzene	1.757	1.674	4.7	92	0.00
88 T	1,4-Dichlorobenzene	1.779	1.669	6.2	91	0.00
89 T	n-Butylbenzene	2.848	2.913	-2.3	92	0.00
90 T	Hexachloroethane	0.638	0.613	3.9	91	0.00
91 T	1,2-Dichlorobenzene	1.696	1.649	2.8	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.227	0.254	-11.9	102	0.00
93 T	1,2,4-Trichlorobenzene	0.910	0.832	8.6	85	0.00
94 T	Hexachlorobutadiene	0.386	0.356	7.8	91	0.00
95 T	Naphthalene	2.769	2.865	-3.5	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091824\
Data File : VN083975.D
Acq On : 18 Sep 2024 11:55
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Sep 19 01:29:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.870	0.814	6.4	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6