

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091918\
 Data File : VN051313.D
 Acq On : 19 Sep 2018 12:47
 Operator : MD\SY
 Sample : J4907-17
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-09-13-18

Quant Time: Sep 19 16:03:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N082818W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 29 01:38:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	83349	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	422233	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	336663	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	153332	28.83	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	96.10%
60) 4-Bromofluorobenzene	12.40	95	94265	18.83	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	62.77%
63) Toluene-d8	10.09	98	480300	30.59	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	101.97%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091918\
Data File : VN051313.D
Acq On : 19 Sep 2018 12:47
Operator : MD\SY
Sample : J4907-17
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-09-13-18

Quant Time: Sep 19 16:03:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N082818W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 29 01:38:44 2018
Response via : Initial Calibration

