

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091918\
 Data File : VN051325.D
 Acq On : 19 Sep 2018 17:42
 Operator : MD\SY
 Sample : PB112994TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112994TB

Quant Time: Sep 20 02:23:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091718W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 18 06:04:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	560562	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	878352	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	737596	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	231206	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	321399	52.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.52%	
35) Dibromofluoromethane	7.59	113	323379	47.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.76%	
50) Toluene-d8	10.09	98	1169034	46.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.68%	
62) 4-Bromofluorobenzene	12.40	95	311018	38.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.52%	
Target Compounds						
16) Acetone	3.82	43	22260	6.80	ug/l	98
18) Methyl Acetate	4.34	43	9060	1.46	ug/l #	76
20) Methylene Chloride	4.55	84	31521	4.73	ug/l	94
43) Isopropyl Acetate	8.17	43	252797	29.89	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091918\
Data File : VN051325.D
Acq On : 19 Sep 2018 17:42
Operator : MD\SY
Sample : PB112994TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112994TB

Quant Time: Sep 20 02:23:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091718W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 18 06:04:06 2018
Response via : Initial Calibration

