

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091918\
 Data File : VN051327.D
 Acq On : 19 Sep 2018 18:31
 Operator : MD\SY
 Sample : PB112984ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112984ZHE#01

Quant Time: Sep 20 02:30:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091718W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 18 06:04:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	543286	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	866855	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	719113	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	234932	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	315172	53.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.76%	
35) Dibromofluoromethane	7.59	113	323568	48.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.06%	
50) Toluene-d8	10.09	98	1142907	45.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.80%	
62) 4-Bromofluorobenzene	12.40	95	299530	37.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.64%	
Target Compounds						
16) Acetone	3.82	43	20797	6.18	ug/l	Qvalue 100
18) Methyl Acetate	4.34	43	9573	1.67	ug/l	98
20) Methylene Chloride	4.56	84	32386	5.01	ug/l	96
43) Isopropyl Acetate	8.17	43	266446	31.92	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091918\
Data File : VN051327.D
Acq On : 19 Sep 2018 18:31
Operator : MD\SY
Sample : PB112984ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112984ZHE#01

Quant Time: Sep 20 02:30:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091718W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 18 06:04:06 2018
Response via : Initial Calibration

