

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091918\
 Data File : VN051328.D
 Acq On : 19 Sep 2018 18:56
 Operator : MD\SY
 Sample : PB112984ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112984ZHE#02

Quant Time: Sep 20 02:32:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091718W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 18 06:04:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	536838	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	845634	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	689196	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	225382	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	309808	53.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.20%	
35) Dibromofluoromethane	7.59	113	316492	48.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.32%	
50) Toluene-d8	10.09	98	1105502	45.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.02%	
62) 4-Bromofluorobenzene	12.40	95	292107	37.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	75.62%	

Target Compounds						Qvalue
16) Acetone	3.82	43	21470	6.93	ug/l	94
18) Methyl Acetate	4.34	43	12071	2.33	ug/l	96
20) Methylene Chloride	4.56	84	31971	5.01	ug/l	96
43) Isopropyl Acetate	8.17	43	294782	36.20	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091918\
Data File : VN051328.D
Acq On : 19 Sep 2018 18:56
Operator : MD\SY
Sample : PB112984ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112984ZHE#02

Quant Time: Sep 20 02:32:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091718W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 18 06:04:06 2018
Response via : Initial Calibration

