

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091918\
 Data File : VN051330.D
 Acq On : 19 Sep 2018 19:45
 Operator : MD\SY
 Sample : PB112984ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112984ZHE#04

Quant Time: Sep 20 03:06:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091718W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 18 06:04:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	540468	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	871330	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	712268	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	230034	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	315680	53.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.48%	
35) Dibromofluoromethane	7.59	113	326160	48.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.34%	
50) Toluene-d8	10.09	98	1139752	45.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.08%	
62) 4-Bromofluorobenzene	12.40	95	299329	37.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.20%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	20673	6.16	ug/l	99
18) Methyl Acetate	4.33	43	12437	2.41	ug/l	97
20) Methylene Chloride	4.55	84	31858	4.96	ug/l	96
43) Isopropyl Acetate	8.17	43	303194	36.14	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091918\
Data File : VN051330.D
Acq On : 19 Sep 2018 19:45
Operator : MD\SY
Sample : PB112984ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112984ZHE#04

Quant Time: Sep 20 03:06:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091718W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 18 06:04:06 2018
Response via : Initial Calibration

