

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091918\
 Data File : VN051333.D
 Acq On : 19 Sep 2018 20:59
 Operator : MD\SY
 Sample : PB112984ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112984ZHE#07

Quant Time: Sep 20 03:15:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091718W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 18 06:04:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	527429	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	852512	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	699395	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	224060	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	314829	54.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.84%	
35) Dibromofluoromethane	7.59	113	319416	48.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.44%	
50) Toluene-d8	10.09	98	1113729	45.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.96%	
62) 4-Bromofluorobenzene	12.40	95	289034	37.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.22%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	22242	7.89	ug/l	95
18) Methyl Acetate	4.33	43	13430	2.74	ug/l	97
20) Methylene Chloride	4.55	84	30232	4.82	ug/l	97
43) Isopropyl Acetate	8.17	43	288037	35.09	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091918\
Data File : VN051333.D
Acq On : 19 Sep 2018 20:59
Operator : MD\SY
Sample : PB112984ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112984ZHE#07

Quant Time: Sep 20 03:15:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091718W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 18 06:04:06 2018
Response via : Initial Calibration

