


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN091918\
Data File : VN051308.D
Acq On : 19 Sep 2018 9:58
Operator : MD\SY
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N082818W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Wed Aug 29 01:38:44 2018



AutoFind: Scans 3305, 3306, 3307; Background Corrected with Scan 3292

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.5	7762	PASS
75	95	30	60	46.7	23400	PASS
95	95	100	100	100.0	50074	PASS
96	95	5	9	7.0	3481	PASS
173	174	0.00	2	0.5	258	PASS
174	95	50	100	97.6	48861	PASS
175	174	5	9	7.2	3511	PASS
176	174	95	101	95.9	46872	PASS
177	176	5	9	6.7	3144	PASS