

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091919\
 Data File : VN058232.D
 Acq On : 20 Sep 2019 1:53
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Sep 20 04:30:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	482462	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	823210	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	753314	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	363249	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.01	65	360342	53.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.24%	
35) Dibromofluoromethane	7.57	113	261645	52.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.56%	
50) Toluene-d8	10.09	98	1022121	52.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.84%	
62) 4-Bromofluorobenzene	12.41	95	376364	51.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.92%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.83	85	186179	46.976	ug/l	99
3) Chloromethane	2.03	50	156668	51.310	ug/l	97
4) Vinyl Chloride	2.17	62	168715	50.513	ug/l	100
5) Bromomethane	2.54	94	86295	48.815	ug/l	99
6) Chloroethane	2.68	64	109998	51.454	ug/l	100
7) Trichlorofluoromethane	3.00	101	249170	46.060	ug/l	99
8) Diethyl Ether	3.39	74	122846	49.560	ug/l	95
9) 1,1,2-Trichlorotrifluoroet	3.74	101	183399	47.021	ug/l	100
10) Methyl Iodide	3.93	142	140814	43.085	ug/l	98
11) Tert butyl alcohol	4.75	59	223401	279.485	ug/l	100
12) 1,1-Dichloroethene	3.72	96	149641	50.559	ug/l	98
13) Acrolein	3.59	56	31793	268.391	ug/l	97
14) Allyl chloride	4.30	41	276343	43.299	ug/l	89
15) Acrylonitrile	4.96	53	584510	274.464	ug/l	99
16) Acetone	3.79	43	497489	218.106	ug/l	98
17) Carbon Disulfide	4.03	76	196082	47.079	ug/l	97
18) Methyl Acetate	4.30	43	316500	52.733	ug/l	100
19) Methyl tert-butyl Ether	5.02	73	783070	53.230	ug/l	99
20) Methylene Chloride	4.53	84	213283	53.423	ug/l	99
21) trans-1,2-Dichloroethene	5.02	96	162074	50.251	ug/l	97
22) Diisopropyl ether	5.93	45	850354	52.451	ug/l	98
23) Vinyl Acetate	5.87	43	3071910	276.914	ug/l	99
24) 1,1-Dichloroethane	5.83	63	442596	51.151	ug/l	99
25) 2-Butanone	6.82	43	806449	259.515	ug/l	99
26) 2,2-Dichloropropane	6.81	77	285107	38.839	ug/l	99
27) cis-1,2-Dichloroethene	6.81	96	247254	50.537	ug/l	99
28) Bromochloromethane	7.18	49	221003	53.991	ug/l	100
29) Tetrahydrofuran	7.19	42	500229	261.332	ug/l	99
30) Chloroform	7.36	83	475437	51.046	ug/l	99
31) Cyclohexane	7.64	56	276799	49.463	ug/l	100
32) 1,1,1-Trichloroethane	7.55	97	370710	52.283	ug/l	100
36) 1,1-Dichloropropene	7.78	75	272869	45.964	ug/l	99
37) Ethyl Acetate	6.91	43	320375	52.002	ug/l	100
38) Carbon Tetrachloride	7.76	117	290659	48.706	ug/l	97

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091919\
 Data File : VN058232.D
 Acq On : 20 Sep 2019 1:53
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleID :
 VSTDCCC050EC

Quant Time: Sep 20 04:30:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	9.07	83	268706	46.818	ug/l	97
40) Benzene	8.03	78	885468	48.834	ug/l	99
41) Methacrylonitrile	7.15	41	150359	48.474	ug/l #	100
42) 1,2-Dichloroethane	8.11	62	381567	50.151	ug/l	100
43) Isopropyl Acetate	8.15	43	576923	50.141	ug/l #	92
44) Trichloroethene	8.82	130	217821	49.090	ug/l	100
45) 1,2-Dichloropropane	9.11	63	279407	51.599	ug/l	99
46) Dibromomethane	9.20	93	166616	50.609	ug/l	100
47) Bromodichloromethane	9.39	83	367669	53.308	ug/l	98
48) Methyl methacrylate	9.19	41	278064	53.226	ug/l	98
49) 1,4-Dioxane	9.19	88	91600	1067.946	ug/l	98
51) 4-Methyl-2-Pentanone	9.98	43	1773695	290.222	ug/l	99
52) Toluene	10.15	92	556319	48.956	ug/l	98
53) t-1,3-Dichloropropene	10.38	75	369458	52.293	ug/l	99
54) cis-1,3-Dichloropropene	9.83	75	392627	50.335	ug/l	98
55) 1,1,2-Trichloroethane	10.56	97	265042	52.557	ug/l	99
56) Ethyl methacrylate	10.43	69	394181	54.175	ug/l	99
57) 1,3-Dichloropropane	10.71	76	449818	51.886	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.69	63	982087	283.365	ug/l	99
59) 2-Hexanone	10.75	43	1272738	280.403	ug/l	99
60) Dibromochloromethane	10.90	129	266432	54.917	ug/l	99
61) 1,2-Dibromoethane	11.00	107	243037	52.004	ug/l	100
64) Tetrachloroethene	10.63	164	175549	47.805	ug/l	98
65) Chlorobenzene	11.43	112	654073	49.363	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.51	131	253894	52.844	ug/l	99
67) Ethyl Benzene	11.51	91	1169327	51.503	ug/l	99
68) m/p-Xylenes	11.62	106	856734	100.190	ug/l	98
69) o-Xylene	11.95	106	428382	50.948	ug/l	99
70) Styrene	11.97	104	772186	53.301	ug/l	99
71) Bromoform	12.13	173	168372	47.464	ug/l #	100
73) Isopropylbenzene	12.25	105	1226924	53.961	ug/l	100
74) N-amyl acetate	12.07	43	546441	57.448	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.51	83	405420	53.246	ug/l	99
76) 1,2,3-Trichloropropane	12.56	75	484043	74.457	ug/l #	100
77) Bromobenzene	12.53	156	285523	49.922	ug/l	98
78) n-propylbenzene	12.60	91	1413849	54.721	ug/l	100
79) 2-Chlorotoluene	12.68	91	856661	53.052	ug/l	100
80) 1,3,5-Trimethylbenzene	12.74	105	1039791	54.080	ug/l	99
81) trans-1,4-Dichloro-2-buten	12.30	75	99414	46.032	ug/l	98
82) 4-Chlorotoluene	12.78	91	889863	52.426	ug/l	100
83) tert-Butylbenzene	13.00	119	917905	53.534	ug/l	99
84) 1,2,4-Trimethylbenzene	13.04	105	1058708	55.332	ug/l	99
85) sec-Butylbenzene	13.18	105	1243555	55.463	ug/l	100
86) p-Isopropyltoluene	13.29	119	1111390	55.756	ug/l	100
87) 1,3-Dichlorobenzene	13.29	146	553010	52.938	ug/l	99
88) 1,4-Dichlorobenzene	13.37	146	553264	51.878	ug/l	100
89) n-Butylbenzene	13.62	91	1018959	53.919	ug/l	100
90) Hexachloroethane	13.88	117	173467	49.648	ug/l	98
91) 1,2-Dichlorobenzene	13.66	146	563724	53.797	ug/l	99
92) 1,2-Dibromo-3-Chloropropan	14.28	75	77558	56.580	ug/l	98

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091919\
Data File : VN058232.D
Acq On : 20 Sep 2019 1:53
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Quant Time: Sep 20 04:30:05 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	14.92	180	363050	54.226	ug/l	99
94) Hexachlorobutadiene	15.01	225	162963	50.360	ug/l	97
95) Naphthalene	15.13	128	1029658	59.343	ug/l	100
96) 1,2,3-Trichlorobenzene	15.30	180	350547	55.396	ug/l	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091919\
Data File : VN058232.D
Acq On : 20 Sep 2019 1:53
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Quant Time: Sep 20 04:30:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.65 min Scan# 1828

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :


VSTDCCC050EC

Tot Ion:168 Resp: 482462

Ion Ratio Lower Upper




168 100

99 60.4 47.4 71.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO



#2

Dichlorodifluoromethane

Concen: 46.976 ug/l

RT: 1.83 min Scan# 19

Delta R.T. 0.00 min

Lab File: VN058232.D


Acq: 20 Sep 2019 1:53

Tgt Ion: 85 Resp: 186179

Ion Ratio Lower Upper



85 100


87 32.0 16.3 48.9



Abundance Ion 84.90 (84.60 to 85.60): VNO

Ion 86.90 (86.60 to 87.60): VNO





#3

Chloromethane

Concen: 51.310 ug/l

RT: 2.03 min Scan# 81

Delta R.T. -0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tot Ion: 50 Resp: 156668

Ion Ratio Lower Upper

50 100

52 31.3 26.3 39.5



#4

Vinyl Chloride

Concen: 50.513 ug/l

RT: 2.17 min Scan# 123

Delta R.T. 0.00 min

Lab File: VN058232.D




Acq: 20 Sep 2019 1:53


Tgt Ion: 62 Resp: 168715

Ion Ratio Lower Upper

62 100

64 31.9 25.4 38.2





#5

Bromomethane

Concen: 48.815 ug/l

RT: 2.54 min Scan# 238

Delta R.T. -0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :


VSTDCCC050EC

Tot Ion: 94 Resp: 86295

Ion Ratio Lower Upper

94 100

96 91.0 73.3 109.9



Abundance Ion 93.90 (93.60 to 94.60): VN0

Ion 95.90 (95.60 to 96.60): VN0

2.54

40000

30000



20000

10000

0

Time-->

2.45 2.50 2.55 2.60 2.65



#6

Chloroethane

Concen: 51.454 ug/l

RT: 2.68 min Scan# 283

Delta R.T. 0.00 min

Lab File: VN058232.D


Acq: 20 Sep 2019 1:53

Tgt Ion: 64 Resp: 109998

Ion Ratio Lower Upper

64 100

66 30.8 24.6 37.0



Abundance Ion 63.90 (63.60 to 64.60): VN0

Ion 65.90 (65.60 to 66.60): VN0

2.68

60000

50000

40000

30000


20000


10000

0

Time-->

2.60 2.70 2.80









#7

Trichlorofluoromethane
 Concen: 46.060 ug/l
 RT: 3.00 min Scan# 381
 Delta R.T. 0.00 min
 Lab File: VN058232.D
 Acq: 20 Sep 2019 1:53

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC




Tot Ion:101 Resp: 249170
 Ion Ratio Lower Upper
 101 100
 103 64.7 51.0 76.6




#8

Diethyl Ether
 Concen: 49.560 ug/l
 RT: 3.39 min Scan# 503
 Delta R.T. 0.00 min
 Lab File: VN058232.D
 Acq: 20 Sep 2019 1:53

Tgt Ion: 74 Resp: 122846
 Ion Ratio Lower Upper
 74 100
 45 102.1 48.5 145.5





#9

1,1,2-Trichlorotrifluoroethane

Concen: 47.021 ug/l

RT: 3.74 min Scan# 611

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC


Tot Ion:101 Resp: 183399

Ion Ratio Lower Upper

101 100

85 46.8 37.3 55.9




151 74.7 59.6 89.4



Abundance Ion 100.90 (100.60 to 101.60): V

Ion 84.90 (84.60 to 85.60): VN

Ion 150.80 (150.50 to 151.50): V



#10

Methyl Iodide

Concen: 43.085 ug/l

RT: 3.93 min Scan# 671

Delta R.T. -0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53


Tgt Ion:142 Resp: 140814

Ion Ratio Lower Upper

142 100

127 45.6 37.5 56.3



141 13.4 11.4 17.2




Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V









#11

Tert butyl alcohol
 Concen: 279.485 ug/l
 RT: 4.75 min Scan# 927
 Delta R.T. 0.00 min
 Lab File: VN058232.D
 Acq: 20 Sep 2019 1:53

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC




Tot Ion: 59 Resp: 223401
 Ion Ratio Lower Upper
 59 100
 57 10.8 8.6 13.0




#12

1,1-Dichloroethene
 Concen: 50.559 ug/l
 RT: 3.72 min Scan# 605
 Delta R.T. 0.00 min
 Lab File: VN058232.D
 Acq: 20 Sep 2019 1:53

Tgt Ion: 96 Resp: 149641
 Ion Ratio Lower Upper
 96 100
 61 183.1 149.5 224.3
 98 64.0 52.4 78.6





#13

Acrolein

Concen: 268.391 ug/l

RT: 3.59 min Scan# 564

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tot Ion: 56 Resp: 31793

Ion Ratio Lower Upper

56 100

55 67.4 56.1 84.1



#14

Allyl chloride

Concen: 43.299 ug/l

RT: 4.30 min Scan# 786

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53




Tgt Ion: 41 Resp: 276343


Ion Ratio Lower Upper

41 100

39 84.8 59.1 88.7

76 35.6 25.1 37.7





#15

Acrylonitrile

Concen: 274.464 ug/l

RT: 4.96 min Scan# 991

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :


MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tot Ion: 53 Resp: 584510

Ion	Ratio	Lower	Upper
53	100		
52	84.1	66.6	100.0
51	37.3	29.7	44.5






Abundance

Ion 53.00 (52.70 to 53.70): VNO

Ion 52.00 (51.70 to 52.70): VNO

Ion 51.00 (50.70 to 51.70): VNO



#16

Acetone

Concen: 218.106 ug/l

RT: 3.79 min Scan# 628


Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Tgt Ion: 43 Resp: 497489



Ion	Ratio	Lower	Upper
43	100		
58	28.0	23.4	35.2




Abundance

Ion 43.00 (42.70 to 43.70): VNO

Ion 58.00 (57.70 to 58.70): VNO





#17

Carbon Disulfide

Concen: 47.079 ug/l

RT: 4.03 min Scan# 703

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :


VSTDCCC050EC

Tot Ion: 76 Resp: 196082

Ion Ratio Lower Upper




76 100

78 8.5 7.7 11.5



Abundance Ion 75.90 (75.60 to 76.60): VN0

Ion 77.90 (77.60 to 78.60): VN0



#18

Methyl Acetate

Concen: 52.733 ug/l

RT: 4.30 min Scan# 786

Delta R.T. 0.00 min

Lab File: VN058232.D


Acq: 20 Sep 2019 1:53

Tgt Ion: 43 Resp: 316500

Ion Ratio Lower Upper



43 100


74 22.6 18.0 27.0



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 74.00 (73.70 to 74.70): VN0









#19
Methyl tert-butyl Ether
Concen: 53.230 ug/l
RT: 5.02 min Scan# 1009
Delta R.T. 0.00 min
Lab File: VN058232.D
Acq: 20 Sep 2019 1:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tot Ion: 73 Resp: 783070
Ion Ratio Lower Upper
73 100
57 22.0 18.1 27.1



#20
Methylene Chloride
Concen: 53.423 ug/l
RT: 4.53 min Scan# 856
Delta R.T. -0.00 min
Lab File: VN058232.D
Acq: 20 Sep 2019 1:53

Tgt Ion: 84 Resp: 213283
Ion Ratio Lower Upper
84 100
49 133.0 107.5 161.3
51 41.3 33.9 50.9
86 65.9 52.4 78.6









#21
trans-1,2-Dichloroethene
Concen: 50.251 ug/l
RT: 5.02 min Scan# 1009
Delta R.T. 0.00 min
Lab File: VN058232.D
Acq: 20 Sep 2019 1:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tot Ion: 96 Resp: 162074
Ion Ratio Lower Upper
96 100
61 155.5 122.2 183.4
98 66.0 49.9 74.9



#22
Diisopropyl ether
Concen: 52.451 ug/l
RT: 5.93 min Scan# 1292
Delta R.T. -0.00 min
Lab File: VN058232.D
Acq: 20 Sep 2019 1:53

Tgt Ion: 45 Resp: 850354
Ion Ratio Lower Upper
45 100
43 64.0 49.7 74.5
87 26.6 20.7 31.1
59 10.8 9.1 13.7









#23
 Vinyl Acetate
 Concen: 276.914 ug/l
 RT: 5.87 min Scan# 1275
 Delta R.T. 0.00 min
 Lab File: VN058232.D
 Acq: 20 Sep 2019 1:53




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tot Ion: 43 Resp: 3071910
 Ion Ratio Lower Upper
 43 100
 86 9.1 7.4 11.2



#24
 1,1-Dichloroethane
 Concen: 51.151 ug/l
 RT: 5.83 min Scan# 1261
 Delta R.T. 0.00 min
 Lab File: VN058232.D
 Acq: 20 Sep 2019 1:53

Tgt Ion: 63 Resp: 442596
 Ion Ratio Lower Upper
 63 100
 98 5.7 2.9 8.6
 100 4.0 1.8 5.3





#25

2-Butanone

Concen: 259.515 ug/l

RT: 6.82 min Scan# 1568

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :


VSTDCCC050EC

Tot Ion: 43 Resp: 806449

Ion Ratio Lower Upper




43 100

72 24.9 19.5 29.3



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 72.00 (71.70 to 72.70): VN0



#26

2,2-Dichloropropane

Concen: 38.839 ug/l

RT: 6.81 min Scan# 1565

Delta R.T. 0.00 min

Lab File: VN058232.D


Acq: 20 Sep 2019 1:53

Tgt Ion: 77 Resp: 285107

Ion Ratio Lower Upper



77 100


97 21.6 10.5 31.6



Abundance Ion 76.90 (76.60 to 77.60): VN0

Ion 96.90 (96.60 to 97.60): VN0





#27

cis-1,2-Dichloroethene

Concen: 50.537 ug/l

RT: 6.81 min Scan# 1566

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC


Tot Ion: 96 Resp: 247254

Ion Ratio Lower Upper

96 100

61 158.5 0.0 319.0




98 64.4 0.0 126.6



Abundance Ion 95.90 (95.60 to 96.60): VN058155.D (-1543) (-)

Ion 60.90 (60.60 to 61.60): VN058155.D (-1543) (-)

Ion 97.90 (97.60 to 98.60): VN058155.D (-1543) (-)



#28

Bromochloromethane

Concen: 53.991 ug/l

RT: 7.18 min Scan# 1680

Delta R.T. -0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53


Tgt Ion: 49 Resp: 221003

Ion Ratio Lower Upper

49 100

129 1.2 0.0 1.8



130 69.6 55.4 83.2




Abundance Ion 48.90 (48.60 to 49.60): VN058155.D (-1657) (-)

Ion 128.80 (128.50 to 129.50): VN058155.D (-1657) (-)

Ion 129.80 (129.50 to 130.50): VN058155.D (-1657) (-)









#29
Tetrahydrofuran
Concen: 261.332 ug/l
RT: 7.19 min Scan# 1685
Delta R.T. 0.00 min
Lab File: VN058232.D
Acq: 20 Sep 2019 1:53

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tot Ion: 42 Resp: 500229
Ion Ratio Lower Upper
42 100
72 41.5 33.8 50.6
71 40.0 31.4 47.0




Abundance Ion 42.00 (41.70 to 42.70): VN0
Ion 72.00 (71.70 to 72.70): VN0
Ion 71.00 (70.70 to 71.70): VN0






#30
Chloroform
Concen: 51.046 ug/l
RT: 7.36 min Scan# 1736
Delta R.T. 0.00 min
Lab File: VN058232.D
Acq: 20 Sep 2019 1:53

Tgt Ion: 83 Resp: 475437
Ion Ratio Lower Upper
83 100
85 63.7 51.4 77.2



Abundance Ion 82.90 (82.60 to 83.60): VN0
Ion 84.90 (84.60 to 85.60): VN0





#31

Cyclohexane

Concen: 49.463 ug/l

RT: 7.64 min Scan# 1824

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC


Tot Ion: 56 Resp: 276799

Ion Ratio Lower Upper

56 100

69 34.4 27.3 40.9




84 80.8 65.0 97.4



Abundance Ion 56.00 (55.70 to 56.70): VN0

Ion 69.00 (68.70 to 69.70): VN0

Ion 84.00 (83.70 to 84.70): VN0



#32

1,1,1-Trichloroethane

Concen: 52.283 ug/l

RT: 7.55 min Scan# 1798

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53


Tgt Ion: 97 Resp: 370710

Ion Ratio Lower Upper

97 100

99 63.1 50.9 76.3



61 48.9 39.2 58.8




Abundance Ion 96.90 (96.60 to 97.60): VN0

Ion 98.90 (98.60 to 99.60): VN0

Ion 60.90 (60.60 to 61.60): VN0





#33

1,2-Dichloroethane-d4

Concen: 53.624 ug/l

RT: 8.01 min Scan# 1940

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :


VSTDCCC050EC

Tot Ion: 65 Resp: 360342

Ion Ratio Lower Upper

65 100

67 51.6 0.0 103.4



Abundance Ion 64.90 (64.60 to 65.60): VN058155.D (-1921) (-)

Ion 66.90 (66.60 to 67.60): VN058155.D (-1921) (-)

8.01

150000



100000

50000

0

7.90 7.95 8.00 8.05 8.10

Time-->



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.57 min Scan# 2115

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53


Tgt Ion: 114 Resp: 823210

Ion Ratio Lower Upper

114 100

63 22.2 0.0 44.2

88 15.8 0.0 31.6



Abundance Ion 113.90 (113.60 to 114.60): VN058155.D (-2097) (-)

Ion 63.00 (62.70 to 63.70): VN058155.D (-2097) (-)

Ion 87.90 (87.60 to 88.60): VN058155.D (-2097) (-)

8.57

500000

400000

300000


200000


100000

0

8.50 8.55 8.60 8.65 8.70

Time-->









#35
 Dibromofluoromethane
 Concen: 52.283 ug/l
 RT: 7.57 min Scan# 1804
 Delta R.T. -0.00 min
 Lab File: VN058232.D
 Acq: 20 Sep 2019 1:53

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
113	100		
111	103.4	81.8	122.6
192	18.2	14.5	21.7




Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V






#36
 1,1-Dichloropropene
 Concen: 45.964 ug/l
 RT: 7.78 min Scan# 1867
 Delta R.T. 0.00 min
 Lab File: VN058232.D
 Acq: 20 Sep 2019 1:53

Tgt Ion	Ratio	Lower	Upper
75	100		
110	33.2	16.6	49.7
77	31.0	24.3	36.5



Abundance Ion 74.90 (74.60 to 75.60): V
 Ion 109.90 (109.60 to 110.60): V
 Ion 76.90 (76.60 to 77.60): V









#37
Ethyl Acetate
Concen: 52.002 ug/l
RT: 6.91 min Scan# 1598
Delta R.T. 0.00 min
Lab File: VN058232.D
Acq: 20 Sep 2019 1:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tot Ion: 43 Resp: 320375
Ion Ratio Lower Upper
43 100
61 13.5 10.7 16.1
70 9.2 7.6 11.4



#38
Carbon Tetrachloride
Concen: 48.706 ug/l
RT: 7.76 min Scan# 1861
Delta R.T. 0.00 min
Lab File: VN058232.D
Acq: 20 Sep 2019 1:53

Tgt Ion: 117 Resp: 290659
Ion Ratio Lower Upper
117 100
119 94.3 78.6 117.8
121 29.7 24.1 36.1









#39
Methylvlcyclohexane
Concen: 46.818 ug/l
RT: 9.07 min Scan# 2270
Delta R.T. 0.00 min
Lab File: VN058232.D
Acq: 20 Sep 2019 1:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tot Ion: 83 Resp: 268706
Ion Ratio Lower Upper
83 100
55 79.0 61.9 92.9
98 41.7 35.4 53.2



#40
Benzene
Concen: 48.834 ug/l
RT: 8.03 min Scan# 1945
Delta R.T. 0.00 min
Lab File: VN058232.D
Acq: 20 Sep 2019 1:53

Tgt Ion: 78 Resp: 885468
Ion Ratio Lower Upper
78 100
77 23.0 18.8 28.2









#41
Methacrylonitrile
Concen: 48.474 ug/l
RT: 7.15 min Scan# 1673
Delta R.T. -0.00 min
Lab File: VN058232.D
Acq: 20 Sep 2019 1:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tot Ion:	41	Resp:	150359
Ion	Ratio	Lower	Upper
41	100		
39	56.4	0.0	0.0#
67	90.2	0.0	0.0#
52	37.8	0.0	0.0#



#42
1,2-Dichloroethane
Concen: 50.151 ug/l
RT: 8.11 min Scan# 1971
Delta R.T. 0.00 min
Lab File: VN058232.D
Acq: 20 Sep 2019 1:53

Tgt Ion:	62	Resp:	381567
Ion	Ratio	Lower	Upper
62	100		
98	7.9	0.0	15.6





#43

Isopropyl Acetate

Concen: 50.141 ug/l

RT: 8.15 min Scan# 1983

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :


MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tot Ion: 43 Resp: 576923




Ion	Ratio	Lower	Upper
43	100		
61	18.7	19.7	29.5#
87	11.7	9.4	14.2



Abundance Ion 43.00 (42.70 to 43.70): VN058155.D (-1964) (-)

Ion 61.00 (60.70 to 61.70): VN058155.D (-1964) (-)

Ion 87.00 (86.70 to 87.70): VN058155.D (-1964) (-)



#44

Trichloroethene

Concen: 49.090 ug/l

RT: 8.82 min Scan# 2193


Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53



Tgt Ion: 130 Resp: 217821


Ion	Ratio	Lower	Upper
130	100		
95	104.2	0.0	207.8



Abundance Ion 129.80 (129.50 to 130.50): VN058155.D (-2176) (-)

Ion 94.90 (94.60 to 95.60): VN058155.D (-2176) (-)









#45
 1,2-Dichloropropane
 Concen: 51.599 ug/l
 RT: 9.11 min Scan# 2281
 Delta R.T. -0.00 min
 Lab File: VN058232.D
 Acq: 20 Sep 2019 1:53




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tot Ion: 63 Resp: 279407
 Ion Ratio Lower Upper
 63 100
 65 30.5 23.8 35.6



#46
 Dibromomethane
 Concen: 50.609 ug/l
 RT: 9.20 min Scan# 2309
 Delta R.T. -0.00 min
 Lab File: VN058232.D
 Acq: 20 Sep 2019 1:53

Tgt Ion: 93 Resp: 166616
 Ion Ratio Lower Upper
 93 100
 95 83.0 66.8 100.2
 174 92.2 73.8 110.8





#47

Bromodichloromethane

Concen: 53.308 ug/l

RT: 9.39 min Scan# 2370

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC


Tot Ion: 83 Resp: 367669

Ion Ratio Lower Upper

83 100

85 64.0 50.1 75.1




127 7.9 6.4 9.6



Abundance Ion 82.90 (82.60 to 83.60): VN058232.D (-2277) (-)

Ion 84.90 (84.60 to 85.60): VN058232.D (-2277) (-)

Ion 126.80 (126.50 to 127.50): VN058232.D (-2277) (-)



#48

Methyl methacrylate

Concen: 53.226 ug/l

RT: 9.19 min Scan# 2307

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53


Tgt Ion: 41 Resp: 278064

Ion Ratio Lower Upper

41 100

69 79.4 62.5 93.7



39 63.9 52.3 78.5




Abundance Ion 41.00 (40.70 to 41.70): VN058232.D (-2214) (-)

Ion 69.00 (68.70 to 69.70): VN058232.D (-2214) (-)

Ion 39.00 (38.70 to 39.70): VN058232.D (-2214) (-)









#49
 1,4-Dioxane
 Concen: 1067.946 ug/l
 RT: 9.19 min Scan# 2307
 Delta R.T. 0.00 min
 Lab File: VN058232.D
 Acq: 20 Sep 2019 1:53

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Tot Ion: 88 Resp: 91600
 Ion Ratio Lower Upper
 88 100
 43 35.7 27.8 41.8
 58 70.2 55.0 82.4




Abundance Ion 88.00 (87.70 to 88.70): VN0
 Ion 43.00 (42.70 to 43.70): VN0
 Ion 58.00 (57.70 to 58.70): VN0






#50
 Toluene-d8
 Concen: 52.421 ug/l
 RT: 10.09 min Scan# 2585
 Delta R.T. 0.00 min
 Lab File: VN058232.D
 Acq: 20 Sep 2019 1:53

Tgt Ion: 98 Resp: 1022121
 Ion Ratio Lower Upper
 98 100
 100 63.5 51.1 76.7



Abundance Ion 98.00 (97.70 to 98.70): VN0
 Ion 100.00 (99.70 to 100.70): VN0





#51

4-Methyl-2-Pentanone

Concen: 290.222 ug/l

RT: 9.98 min Scan# 2552

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :


VSTDCCC050EC

Tot Ion: 43 Resp: 1773695

Ion Ratio Lower Upper




43 100

58 37.3 30.2 45.4



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0



#52

Toluene

Concen: 48.956 ug/l

RT: 10.15 min Scan# 2605

Delta R.T. 0.00 min

Lab File: VN058232.D


Acq: 20 Sep 2019 1:53

Tgt Ion: 92 Resp: 556319

Ion Ratio Lower Upper



92 100


91 178.1 140.2 210.2



Abundance Ion 92.00 (91.70 to 92.70): VN0

Ion 91.00 (90.70 to 91.70): VN0





#53

t-1,3-Dichloropropene

Concen: 52.293 ug/l

RT: 10.38 min Scan# 2676

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tot Ion: 75 Resp: 369458

Ion Ratio Lower Upper

75 100

77 31.7 24.9 37.3



#54

cis-1,3-Dichloropropene

Concen: 50.335 ug/l

RT: 9.83 min Scan# 2507

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53




Tgt Ion: 75 Resp: 392627


Ion Ratio Lower Upper

75 100

77 31.2 25.4 38.0

39 57.8 45.0 67.6





#55

1,1,2-Trichloroethane

Concen: 52.557 ug/l

RT: 10.56 min Scan# 2732

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tot Ion: 97 Resp: 265042


Ion Ratio Lower Upper

97 100

83 87.2 70.4 105.6

85 56.8 46.8 70.2

99 61.7 49.9 74.9






Abundance Ion 96.90 (96.60 to 97.60): VN0

Ion 82.90 (82.60 to 83.60): VN0

Ion 84.90 (84.60 to 85.60): VN0

Ion 98.90 (98.60 to 99.60): VN0



#56

Ethyl methacrylate

Concen: 54.175 ug/l

RT: 10.43 min Scan# 2691

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53


Tgt Ion: 69 Resp: 394181

Ion Ratio Lower Upper

69 100

41 72.0 57.5 86.3



39 43.9 36.1 54.1




Abundance Ion 69.00 (68.70 to 69.70): VN0

Ion 41.00 (40.70 to 41.70): VN0

Ion 39.00 (38.70 to 39.70): VN0





#57

1,3-Dichloropropane

Concen: 51.886 ug/l

RT: 10.71 min Scan# 2778

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tot Ion: 76 Resp: 449818

Ion Ratio Lower Upper

76 100

78 32.1 25.4 38.0



#58

2-Chloroethyl Vinyl ether

Concen: 283.365 ug/l

RT: 9.69 min Scan# 2462

Delta R.T. 0.00 min

Lab File: VN058232.D




Acq: 20 Sep 2019 1:53


Tgt Ion: 63 Resp: 982087

Ion Ratio Lower Upper

63 100

106 25.3 19.9 29.9









#59
2-Hexanone
Concen: 280.403 ug/l
RT: 10.75 min Scan# 2792
Delta R.T. 0.00 min
Lab File: VN058232.D
Acq: 20 Sep 2019 1:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tot Ion: 43 Resp: 1272738
Ion Ratio Lower Upper
43 100
58 51.1 26.0 78.0



#60
Dibromochloromethane
Concen: 54.917 ug/l
RT: 10.90 min Scan# 2839
Delta R.T. 0.00 min
Lab File: VN058232.D
Acq: 20 Sep 2019 1:53

Tgt Ion: 129 Resp: 266432
Ion Ratio Lower Upper
129 100
127 76.2 38.7 116.1









#61
 1,2-Dibromoethane
 Concen: 52.004 ug/l
 RT: 11.00 min Scan# 2871
 Delta R.T. 0.00 min
 Lab File: VN058232.D
 Acq: 20 Sep 2019 1:53




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tot Ion: 107 Resp: 243037
 Ion Ratio Lower Upper
 107 100
 109 94.0 75.4 113.2



#62
 4-Bromofluorobenzene
 Concen: 51.963 ug/l
 RT: 12.41 min Scan# 3307
 Delta R.T. 0.00 min
 Lab File: VN058232.D
 Acq: 20 Sep 2019 1:53

Tgt Ion: 95 Resp: 376364
 Ion Ratio Lower Upper
 95 100
 174 74.9 0.0 152.2
 176 72.6 0.0 148.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. 0.00 min
Lab File: VN058232.D
Acq: 20 Sep 2019 1:53

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tot Ion:117 Resp: 753314
Ion Ratio Lower Upper
117 100
82 59.3 46.9 70.3
119 31.2 25.3 37.9




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V








#64
Tetrachloroethene
Concen: 47.805 ug/l
RT: 10.63 min Scan# 2754
Delta R.T. 0.00 min
Lab File: VN058232.D
Acq: 20 Sep 2019 1:53

Tgt Ion:164 Resp: 175549
Ion Ratio Lower Upper
164 100
166 123.4 102.2 153.4
129 99.5 79.6 119.4
131 95.1 76.0 114.0



Abundance Ion 163.80 (163.50 to 164.50): V
Ion 165.80 (165.50 to 166.50): V
Ion 128.80 (128.50 to 129.50): V
Ion 130.80 (130.50 to 131.50): V









#65
Chlorobenzene
Concen: 49.363 ug/l
RT: 11.43 min Scan# 3004
Delta R.T. 0.00 min
Lab File: VN058232.D
Acq: 20 Sep 2019 1:53


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tot Ion:112 Resp: 654073
Ion Ratio Lower Upper
112 100
114 31.3 25.1 37.7



#66
1,1,1,2-Tetrachloroethane
Concen: 52.844 ug/l
RT: 11.51 min Scan# 3028
Delta R.T. 0.00 min
Lab File: VN058232.D
Acq: 20 Sep 2019 1:53

Tgt Ion:131 Resp: 253894
Ion Ratio Lower Upper
131 100
133 96.6 47.8 143.3
119 66.7 33.1 99.3









#67
Ethyl Benzene
Concen: 51.503 ug/l
RT: 11.51 min Scan# 3029
Delta R.T. 0.00 min
Lab File: VN058232.D
Acq: 20 Sep 2019 1:53




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tot Ion: 91 Resp: 1169327
Ion Ratio Lower Upper
91 100
106 30.8 24.0 36.0



#68
m/p-Xylenes
Concen: 100.190 ug/l
RT: 11.62 min Scan# 3063
Delta R.T. 0.00 min
Lab File: VN058232.D
Acq: 20 Sep 2019 1:53

Tgt Ion: 106 Resp: 856734
Ion Ratio Lower Upper
106 100
91 207.3 163.6 245.4









#69
o-Xylene
Concen: 50.948 ug/l
RT: 11.95 min Scan# 3165
Delta R.T. 0.00 min
Lab File: VN058232.D
Acq: 20 Sep 2019 1:53

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tot Ion:106 Resp: 428382
Ion Ratio Lower Upper
106 100
91 221.9 111.8 335.3




Abundance Ion 106.00 (105.70 to 106.70): V
Ion 91.00 (90.70 to 91.70): VN






#70
Styrene
Concen: 53.301 ug/l
RT: 11.97 min Scan# 3170
Delta R.T. 0.00 min
Lab File: VN058232.D
Acq: 20 Sep 2019 1:53

Tgt Ion:104 Resp: 772186
Ion Ratio Lower Upper
104 100
78 52.1 41.8 62.8
103 54.5 44.2 66.2



Abundance Ion 104.00 (103.70 to 104.70): V
Ion 78.00 (77.70 to 78.70): VN
Ion 103.00 (102.70 to 103.70): V





#71

Bromoform

Concen: 47.464 ug/l

RT: 12.13 min Scan# 3221

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC


Tot Ion:173 Resp: 168372

Ion Ratio Lower Upper




173 100

175 48.7 24.2 72.6

254 0.0 0.1 0.1#



Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3600

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53


Tgt Ion:152 Resp: 363249

Ion Ratio Lower Upper



152 100


115 60.6 30.1 90.3

150 174.4 0.0 346.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: 53.961 ug/l

RT: 12.25 min Scan# 3259

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :


VSTDCCC050EC

Tot Ion:105 Resp: 1226924

Ion Ratio Lower Upper




105 100

120 25.5 12.8 38.3



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#74

N-ethyl acetate

Concen: 57.448 ug/l

RT: 12.07 min Scan# 3203

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Tgt Ion: 43 Resp: 546441


Ion Ratio Lower Upper

43 100

70 38.8 31.0 46.6

55 23.7 18.5 27.7

61 22.1 18.2 27.2






Abundance Ion 43.00 (42.70 to 43.70): V

Ion 70.00 (69.70 to 70.70): V

Ion 55.00 (54.70 to 55.70): V

Ion 61.00 (60.70 to 61.70): V





#75

1,1,2,2-Tetrachloroethane

Concen: 53.246 ug/l

RT: 12.51 min Scan# 3339

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :


MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tot Ion: 83 Resp: 405420



Ion	Ratio	Lower	Upper
83	100		
131	9.3	4.8	14.3
85	64.2	31.9	95.5



Abundance Ion 82.90 (82.60 to 83.60): VN0

Ion 130.80 (130.50 to 131.50): VN0

Ion 84.90 (84.60 to 85.60): VN0



#76

1,2,3-Trichloropropane

Concen: 74.457 ug/l

RT: 12.56 min Scan# 3354

Delta R.T. 0.00 min

Lab File: VN058232.D



Acq: 20 Sep 2019 1:53


Tgt Ion: 75 Resp: 484043

Ion	Ratio	Lower	Upper
75	100		
77	0.0	0.0	0.0

Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 77.00 (76.70 to 77.70): VN0





#77

Bromobenzene

Concen: 49.922 ug/l

RT: 12.53 min Scan# 3345

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC


Tot Ion:156 Resp: 285523

Ion Ratio Lower Upper

156 100

77 226.8 111.7 335.1




158 97.7 47.9 143.8



Abundance Ion 155.80 (155.50 to 156.50): V

Ion 77.00 (76.70 to 77.70): VN0

Ion 157.80 (157.50 to 158.50): V



#78

n-propylbenzene

Concen: 54.721 ug/l

RT: 12.60 min Scan# 3366

Delta R.T. 0.00 min

Lab File: VN058232.D


Acq: 20 Sep 2019 1:53

Tgt Ion: 91 Resp: 1413849

Ion Ratio Lower Upper



91 100


120 22.1 11.1 33.3



Abundance Ion 91.00 (90.70 to 91.70): VN0

Ion 120.00 (119.70 to 120.70): V





#79

2-Chlorotoluene

Concen: 53.052 ug/l

RT: 12.68 min Scan# 3391

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tot Ion: 91 Resp: 856661

Ion Ratio Lower Upper

91 100

126 32.5 16.4 49.1



#80

1,3,5-Trimethylbenzene

Concen: 54.080 ug/l

RT: 12.74 min Scan# 3410

Delta R.T. 0.00 min

Lab File: VN058232.D




Acq: 20 Sep 2019 1:53


Tgt Ion: 105 Resp: 1039791

Ion Ratio Lower Upper

105 100

120 47.6 23.4 70.0





#81

trans-1,4-Dichloro-2-butene

Concen: 46.032 ug/l

RT: 12.30 min Scan# 3275

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC


Tot Ion: 75 Resp: 99414

Ion Ratio Lower Upper

75 100

53 110.9 90.1 135.1




89 42.6 36.2 54.2



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 53.00 (52.70 to 53.70): VN0

Ion 88.90 (88.60 to 89.60): VN0



#82

4-Chlorotoluene

Concen: 52.426 ug/l

RT: 12.78 min Scan# 3422

Delta R.T. 0.00 min

Lab File: VN058232.D


Acq: 20 Sep 2019 1:53

Tgt Ion: 91 Resp: 889863

Ion Ratio Lower Upper



91 100


126 31.6 15.7 47.1



Abundance Ion 91.00 (90.70 to 91.70): VN0

Ion 125.90 (125.60 to 126.60): V









#83
 tert-Butylbenzene
 Concen: 53.534 ug/l
 RT: 13.00 min Scan# 3491
 Delta R.T. 0.00 min
 Lab File: VN058232.D
 Acq: 20 Sep 2019 1:53

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Tot Ion:119 Resp: 917905
 Ion Ratio Lower Upper
 119 100
 91 68.6 33.8 101.3
 134 23.1 11.6 34.8




Abundance Ion 119.00 (118.70 to 119.70): V
 Ion 91.00 (90.70 to 91.70): V
 Ion 134.00 (133.70 to 134.70): V






#84
 1,2,4-Trimethylbenzene
 Concen: 55.332 ug/l
 RT: 13.04 min Scan# 3505
 Delta R.T. 0.00 min
 Lab File: VN058232.D
 Acq: 20 Sep 2019 1:53

Tgt Ion:105 Resp: 1058708
 Ion Ratio Lower Upper
 105 100
 120 43.8 22.1 66.5



Abundance Ion 105.00 (104.70 to 105.70): V
 Ion 120.00 (119.70 to 120.70): V





#85

sec-Butylbenzene

Concen: 55.463 ug/l

RT: 13.18 min Scan# 3547

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :


VSTDCCC050EC

Tot Ion:105 Resp: 1243555

Ion Ratio Lower Upper




105 100

134 18.8 9.5 28.5



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V



#86

p-Isopropyltoluene

Concen: 55.756 ug/l

RT: 13.29 min Scan# 3583

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53


Tgt Ion:119 Resp: 1111390

Ion Ratio Lower Upper

119 100

134 25.4 12.7 38.0



91 24.4 12.3 36.8




Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): V









#87
 1,3-Dichlorobenzene
 Concen: 52.938 ug/l
 RT: 13.29 min Scan# 3581
 Delta R.T. 0.00 min
 Lab File: VN058232.D
 Acq: 20 Sep 2019 1:53




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tot Ion:146 Resp: 553010
 Ion Ratio Lower Upper
 146 100
 111 41.2 20.9 62.8
 148 62.8 32.0 96.2



#88
 1,4-Dichlorobenzene
 Concen: 51.878 ug/l
 RT: 13.37 min Scan# 3606
 Delta R.T. 0.00 min
 Lab File: VN058232.D
 Acq: 20 Sep 2019 1:53

Tgt Ion:146 Resp: 553264
 Ion Ratio Lower Upper
 146 100
 111 41.2 20.5 61.5
 148 63.6 31.9 95.5





#89

n-Butylbenzene

Concen: 53.919 ug/l

RT: 13.62 min Scan# 3685

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tot Ion: 91 Resp: 1018959

Ion Ratio Lower Upper

91 100

92 51.4 25.7 77.0

134 24.1 11.8 35.4



#90

Hexachloroethane

Concen: 49.648 ug/l

RT: 13.88 min Scan# 3765

Delta R.T. -0.00 min

Lab File: VN058232.D




Acq: 20 Sep 2019 1:53


Tgt Ion: 117 Resp: 173467

Ion Ratio Lower Upper

117 100

201 76.6 37.5 112.5









#91
 1,2-Dichlorobenzene
 Concen: 53.797 ug/l
 RT: 13.66 min Scan# 3697
 Delta R.T. 0.00 min
 Lab File: VN058232.D
 Acq: 20 Sep 2019 1:53




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tot Ion	Ratio	Lower	Upper
146	100		
111	42.5	21.1	63.1
148	62.5	31.8	95.4



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 56.580 ug/l
 RT: 14.28 min Scan# 3889
 Delta R.T. 0.00 min
 Lab File: VN058232.D
 Acq: 20 Sep 2019 1:53

Tgt Ion	Ratio	Lower	Upper
75	100		
155	75.2	38.3	114.8
157	93.0	48.0	144.0









#93
 1,2,4-Trichlorobenzene
 Concen: 54.226 ug/l
 RT: 14.92 min Scan# 4088
 Delta R.T. 0.00 min
 Lab File: VN058232.D
 Acq: 20 Sep 2019 1:53




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tot Ion:180 Resp: 363050
 Ion Ratio Lower Upper
 180 100
 182 94.5 47.8 143.3
 145 30.1 15.4 46.4



#94
 Hexachlorobutadiene
 Concen: 50.360 ug/l
 RT: 15.01 min Scan# 4118
 Delta R.T. 0.00 min
 Lab File: VN058232.D
 Acq: 20 Sep 2019 1:53

Tgt Ion:225 Resp: 162963
 Ion Ratio Lower Upper
 225 100
 223 62.5 32.6 97.7
 227 64.2 33.0 99.0





#95

Naphthalene

Concen: 59.343 ug/l

RT: 15.13 min Scan# 4155

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC


Tot Ion:128 Resp: 1029658

Ion Ratio Lower Upper

128 100

127 12.8 10.2 15.2




129 10.5 8.6 12.8



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V



#96

1,2,3-Trichlorobenzene

Concen: 55.396 ug/l

RT: 15.30 min Scan# 4207

Delta R.T. 0.00 min

Lab File: VN058232.D

Acq: 20 Sep 2019 1:53


Tgt Ion:180 Resp: 350547

Ion Ratio Lower Upper

180 100

182 94.6 47.4 142.2

145 32.2 16.1 48.2



Abundance Ion 179.80 (179.50 to 180.50): V

Ion 181.80 (181.50 to 182.50): V

Ion 144.90 (144.60 to 145.60): V

