

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091919\
 Data File : VN058191.D
 Acq On : 19 Sep 2019 9:50
 Operator : JC/SP
 Sample : VN0919WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0919WBL01

Quant Time: Sep 20 02:22:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	518690	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	826786	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	708923	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	300044	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	356030	49.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.56%	
35) Dibromofluoromethane	7.57	113	249091	49.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.12%	
50) Toluene-d8	10.08	98	1020287	52.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.20%	
62) 4-Bromofluorobenzene	12.40	95	331538	45.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.16%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091919\
Data File : VN058191.D
Acq On : 19 Sep 2019 9:50
Operator : JC/SP
Sample : VN0919WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0919WBL01

Quant Time: Sep 20 02:22:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

