

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091919\
 Data File : VN058198.D
 Acq On : 19 Sep 2019 12:22
 Operator : JC/SP
 Sample : K4896-11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE104D2-20190912

Quant Time: Sep 20 07:51:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	513722	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	831796	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	710253	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	289701	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	366892	51.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.56%	
35) Dibromofluoromethane	7.58	113	256331	50.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.38%	
50) Toluene-d8	10.08	98	1038718	52.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.44%	
62) 4-Bromofluorobenzene	12.40	95	333409	45.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.12%	
Target Compounds						
16) Acetone	3.80	43	17984	7.405	ug/l	94
27) cis-1,2-Dichloroethene	6.82	96	57174	10.975	ug/l	87
30) Chloroform	7.36	83	10330	1.042	ug/l	91
44) Trichloroethene	8.82	130	289849	64.648	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091919\
Data File : VN058198.D
Acq On : 19 Sep 2019 12:22
Operator : JC/SP
Sample : K4896-11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE104D2-20190912

Quant Time: Sep 20 07:51:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

