

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091919\
 Data File : VN058203.D
 Acq On : 19 Sep 2019 14:31
 Operator : JC/SP
 Sample : K4896-15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP03-20190912

Quant Time: Sep 20 08:01:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	484181	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	790272	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	661126	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	261923	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	343358	50.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.84%	
35) Dibromofluoromethane	7.58	113	238654	49.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.36%	
50) Toluene-d8	10.08	98	970392	51.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.68%	
62) 4-Bromofluorobenzene	12.41	95	302842	43.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.12%	
Target Compounds						
16) Acetone	3.80	43	16610	7.256	ug/l	95
44) Trichloroethene	8.82	130	13639	3.202	ug/l	98
64) Tetrachloroethene	10.63	164	7341	2.278	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091919\
Data File : VN058203.D
Acq On : 19 Sep 2019 14:31
Operator : JC/SP
Sample : K4896-15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP03-20190912

Quant Time: Sep 20 08:01:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

