

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091919\
 Data File : VN058228.D
 Acq On : 20 Sep 2019 00:24
 Operator : JC/SP
 Sample : K4974-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE134D2-20190916

Quant Time: Sep 20 07:42:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	524020	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	849898	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	733167	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	304667	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.01	65	373244	51.14	ug/l	0.00
Spiked Amount			Recovery	=	102.28%	
35) Dibromofluoromethane	7.57	113	257717	49.88	ug/l	0.00
Spiked Amount			Recovery	=	99.76%	
50) Toluene-d8	10.08	98	1061851	52.75	ug/l	0.00
Spiked Amount			Recovery	=	105.50%	
62) 4-Bromofluorobenzene	12.40	95	346577	46.35	ug/l	0.00
Spiked Amount			Recovery	=	92.70%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	3.74	101	69566	16.421	ug/l	98
12) 1,1-Dichloroethene	3.72	96	2024	0.276	ug/l	80
16) Acetone	3.80	43	12581	5.078	ug/l	91
27) cis-1,2-Dichloroethene	6.81	96	11158	2.100	ug/l	87
44) Trichloroethene	8.82	130	629364	137.384	ug/l	98
64) Tetrachloroethene	10.63	164	61660	17.252	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091919\
Data File : VN058228.D
Acq On : 20 Sep 2019 00:24
Operator : JC/SP
Sample : K4974-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE134D2-20190916

Quant Time: Sep 20 07:42:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

