

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091922\
 Data File : VN074577.D
 Acq On : 19 Sep 2022 11:25
 Operator : JC\MD
 Sample : VN0919MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0919MBL01

Quant Time: Sep 20 04:18:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	252054	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	480054	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	467812	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	198967	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	207504	48.520	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.040%
35) Dibromofluoromethane	7.945	113	167055	52.386	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.780%
50) Toluene-d8	10.381	98	608505	49.341	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.680%
62) 4-Bromofluorobenzene	12.675	95	223673	52.954	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091922\
Data File : VN074577.D
Acq On : 19 Sep 2022 11:25
Operator : JC\MD
Sample : VN0919MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0919MBL01

Quant Time: Sep 20 04:18:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

