

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091922\
 Data File : VN074581.D
 Acq On : 19 Sep 2022 13:12
 Operator : JC\MD
 Sample : N4647-01DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1034-MW-01(23)DL

Quant Time: Sep 20 04:19:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	249360	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	464064	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	474127	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	203057	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	196848	46.525	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.060%
35) Dibromofluoromethane	7.951	113	165623	53.727	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.460%
50) Toluene-d8	10.381	98	587687	49.295	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.580%
62) 4-Bromofluorobenzene	12.675	95	218158	53.428	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.860%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	15461	5.711	ug/l	96
64) Tetrachloroethene	10.916	164	161605	49.666	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091922\
Data File : VN074581.D
Acq On : 19 Sep 2022 13:12
Operator : JC\MD
Sample : N4647-01DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1034-MW-01(23)DL

Quant Time: Sep 20 04:19:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

