

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091922\
 Data File : VN074586.D
 Acq On : 19 Sep 2022 15:14
 Operator : JC\MD
 Sample : N4647-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1040-TB090822

Quant Time: Sep 20 04:19:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	252384	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	489890	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	490017	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	199869	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	200691	46.865	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.740%
35) Dibromofluoromethane	7.951	113	164477	50.542	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.080%
50) Toluene-d8	10.381	98	603216	47.930	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.860%
62) 4-Bromofluorobenzene	12.674	95	221982	51.499	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091922\
Data File : VN074586.D
Acq On : 19 Sep 2022 15:14
Operator : JC\MD
Sample : N4647-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1040-TB090822

Quant Time: Sep 20 04:19:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

