

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091922\
 Data File : VN074587.D
 Acq On : 19 Sep 2022 15:39
 Operator : JC\MD
 Sample : N4647-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1039-FB090822

Quant Time: Sep 20 04:20:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	244005	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	472326	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	484263	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	211078	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	197644	47.739	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.480%
35) Dibromofluoromethane	7.951	113	162624	51.831	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.660%
50) Toluene-d8	10.380	98	593587	48.919	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.840%
62) 4-Bromofluorobenzene	12.674	95	226626	54.531	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.060%
Target Compounds						
					Qvalue	
16) Acetone	4.222	43	77735	29.342	ug/l	90
25) 2-Butanone	7.257	43	17164	3.902	ug/l #	68
52) Toluene	10.439	92	10337	1.026	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091922\
Data File : VN074587.D
Acq On : 19 Sep 2022 15:39
Operator : JC\MD
Sample : N4647-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1039-FB090822

Quant Time: Sep 20 04:20:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

