

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091922\
 Data File : VN074589.D
 Acq On : 19 Sep 2022 16:28
 Operator : JC\MD
 Sample : N4471-15DL2 2000X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-06-1.5-11.5-083122DL2

Quant Time: Sep 20 04:20:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	252402	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	486990	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	484006	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	220434	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	198741	46.407	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.820%
35) Dibromofluoromethane	7.951	113	165932	51.293	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.580%
50) Toluene-d8	10.380	98	598656	47.851	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.700%
62) 4-Bromofluorobenzene	12.674	95	233815	54.567	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.140%
Target Compounds						Qvalue
44) Trichloroethene	9.151	130	49722	12.977	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091922\
Data File : VN074589.D
Acq On : 19 Sep 2022 16:28
Operator : JC\MD
Sample : N4471-15DL2 2000X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-06-1.5-11.5-083122DL2

Quant Time: Sep 20 04:20:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

