

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091922\
 Data File : VN074600.D
 Acq On : 19 Sep 2022 20:57
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleID :
 VSTDCCC050

Quant Time: Sep 20 04:22:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	54	0.00
2 T	Dichlorodifluoromethane	0.544	0.485	10.8	57	0.00
3 P	Chloromethane	1.009	1.022	-1.3	66	0.00
4 C	Vinyl Chloride	0.979	1.185	-21.0#	77	0.00
5 T	Bromomethane	0.517	0.809	-56.5#	92	0.00
6 T	Chloroethane	0.669	0.877	-31.1#	83	-0.01
7 T	Trichlorofluoromethane	1.082	1.410	-30.3#	81	-0.01
8 T	Diethyl Ether	0.552	0.558	-1.1	63	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.669	0.662	1.0	61	0.00
10 T	Methyl Iodide	0.789	0.823	-4.3	60	0.00
11 T	Tert butyl alcohol	0.276	0.297	-7.6	69	0.00
12 CM	1,1-Dichloroethene	0.719	0.655	8.9#	61	0.00
13 T	Acrolein	0.101	0.124	-22.8	70	0.00
14 T	Allyl chloride	1.572	1.620	-3.1	63	0.00
15 T	Acrylonitrile	0.621	0.714	-15.0	68	0.00
16 T	Acetone	0.543	0.647	-19.2	75	0.00
17 T	Carbon Disulfide	1.806	1.652	8.5	57	0.00
18 T	Methyl Acetate	1.686	1.895	-12.4	70	0.00
19 T	Methyl tert-butyl Ether	2.775	2.918	-5.2	64	0.00
20 T	Methylene Chloride	0.930	0.852	8.4	62	0.00
21 T	trans-1,2-Dichloroethene	0.785	0.692	11.8	57	0.00
22 T	Diisopropyl ether	3.157	3.516	-11.4	66	0.00
23 T	Vinyl Acetate	2.785	3.205	-15.1	66	0.00
24 P	1,1-Dichloroethane	1.578	1.644	-4.2	64	0.00
25 T	2-Butanone	0.901	1.091	-21.1	72	0.00
26 T	2,2-Dichloropropane	1.361	0.976	28.3#	45#	0.00
27 T	cis-1,2-Dichloroethene	0.926	0.909	1.8	60	0.00
28 T	Bromochloromethane	0.544	0.496	8.8	65	0.00
29 T	Tetrahydrofuran	0.593	0.718	-21.1	71	0.00
30 C	Chloroform	1.457	1.575	-8.1#	64	0.00
31 T	Cyclohexane	1.506	1.474	2.1	61	0.00
32 T	1,1,1-Trichloroethane	1.278	1.277	0.1	62	0.00
33 S	1,2-Dichloroethane-d4	0.848	0.901	-6.3	61	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	56	0.00
35 S	Dibromofluoromethane	0.332	0.358	-7.8	61	0.00
36 T	1,1-Dichloropropene	0.588	0.568	3.4	60	0.00
37 T	Ethyl Acetate	0.908	1.065	-17.3	69	0.00
38 T	Carbon Tetrachloride	0.543	0.538	0.9	60	0.00
39 T	Methylcyclohexane	0.699	0.670	4.1	56	0.00
40 TM	Benzene	1.818	1.861	-2.4	61	0.00
41 T	Methacrylonitrile	0.439	0.533	-21.4	74	0.00
42 TM	1,2-Dichloroethane	0.624	0.675	-8.2	64	0.00
43 T	Isopropyl Acetate	1.422	1.633	-14.8	69	0.00
44 TM	Trichloroethene	0.393	0.369	6.1	59	0.00
45 C	1,2-Dichloropropane	0.500	0.520	-4.0#	63	0.00
46 T	Dibromomethane	0.313	0.314	-0.3	61	0.00
47 T	Bromodichloromethane	0.613	0.635	-3.6	62	0.00
48 T	Methyl methacrylate	0.631	0.684	-8.4	67	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091922\
 Data File : VN074600.D
 Acq On : 19 Sep 2022 20:57
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleID :
 VSTDCCC050

Quant Time: Sep 20 04:22:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.012	0.014	-16.7	69	0.00
50 S	Toluene-d8	1.285	1.313	-2.2	60	0.00
51 T	4-Methyl-2-Pentanone	0.929	1.151	-23.9	72	0.00
52 CM	Toluene	1.066	1.229	-15.3#	67	0.00
53 T	t-1,3-Dichloropropene	0.715	0.706	1.3	58	0.00
54 T	cis-1,3-Dichloropropene	0.765	0.757	1.0	59	0.00
55 T	1,1,2-Trichloroethane	0.452	0.455	-0.7	61	0.00
56 T	Ethyl methacrylate	0.807	0.866	-7.3	63	0.00
57 T	1,3-Dichloropropane	0.792	0.827	-4.4	63	0.00
58 T	2-Chloroethyl Vinyl ether	0.355	0.341	3.9	55	0.00
59 T	2-Hexanone	0.709	0.888	-25.2#	72	0.00
60 T	Dibromochloromethane	0.432	0.453	-4.9	61	0.00
61 T	1,2-Dibromoethane	0.439	0.474	-8.0	63	0.00
62 S	4-Bromofluorobenzene	0.440	0.465	-5.7	58	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	56	0.00
64 T	Tetrachloroethene	0.343	0.315	8.2	57	0.00
65 PM	Chlorobenzene	1.255	1.251	0.3	61	0.00
66 T	1,1,1,2-Tetrachloroethane	0.456	0.459	-0.7	61	0.00
67 C	Ethyl Benzene	2.305	2.442	-5.9#	63	0.00
68 T	m/p-Xylenes	0.852	0.957	-12.3	65	0.00
69 T	o-Xylene	0.869	0.952	-9.6	64	0.00
70 T	Styrene	1.445	1.518	-5.1	61	0.00
71 P	Bromoform	0.335	0.351	-4.8	60	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	56	0.00
73 T	Isopropylbenzene	5.120	5.076	0.9	62	0.00
74 T	N-amyl acetate	3.201	3.471	-8.4	70	0.00
75 P	1,1,2,2-Tetrachloroethane	2.052	2.033	0.9	65	0.00
76 T	1,2,3-Trichloropropane	1.539	1.559	-1.3	66	0.00
77 T	Bromobenzene	1.094	1.069	2.3	62	0.00
78 T	n-propylbenzene	6.098	6.131	-0.5	61	0.00
79 T	2-Chlorotoluene	3.640	3.555	2.3	61	0.00
80 T	1,3,5-Trimethylbenzene	4.173	4.124	1.2	60	0.00
81 T	trans-1,4-Dichloro-2-butene	0.713	0.610	14.4	52	0.00
82 T	4-Chlorotoluene	3.460	3.414	1.3	61	0.00
83 T	tert-Butylbenzene	3.798	3.705	2.4	61	0.00
84 T	1,2,4-Trimethylbenzene	4.200	4.303	-2.5	63	0.00
85 T	sec-Butylbenzene	5.215	5.245	-0.6	60	0.00
86 T	p-Isopropyltoluene	4.066	4.140	-1.8	60	0.00
87 T	1,3-Dichlorobenzene	2.010	1.949	3.0	60	0.00
88 T	1,4-Dichlorobenzene	2.020	1.976	2.2	62	0.00
89 T	n-Butylbenzene	3.665	3.698	-0.9	59	0.00
90 T	Hexachloroethane	0.885	0.801	9.5	56	0.00
91 T	1,2-Dichlorobenzene	2.085	1.998	4.2	61	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.453	0.432	4.6	61	0.00
93 T	1,2,4-Trichlorobenzene	0.993	1.001	-0.8	62	0.00
94 T	Hexachlorobutadiene	0.437	0.404	7.6	60	0.00
95 T	Naphthalene	4.204	4.420	-5.1	64	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091922\
Data File : VN074600.D
Acq On : 19 Sep 2022 20:57
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Sep 20 04:22:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T 1,2,3-Trichlorobenzene	1.032	0.983	4.7	59	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6