


BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091923\
Data File : VN079296.D
Acq On : 19 Sep 2023 10:27
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090723W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Fri Sep 08 05:09:17 2023



AutoFind: Scans 1852, 1853, 1854; Background Corrected with Scan 1845

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.5	6877	PASS
75	95	30	60	52.9	16917	PASS
95	95	100	100	100.0	31963	PASS
96	95	5	9	6.1	1952	PASS
173	174	0.00	2	1.1	265	PASS
174	95	50	100	74.8	23915	PASS
175	174	5	9	7.5	1789	PASS
176	174	95	101	95.5	22837	PASS
177	176	5	9	7.6	1742	PASS