

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091924\
 Data File : VN084036.D
 Acq On : 19 Sep 2024 16:03
 Operator : JC\MD
 Sample : P4085-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-2

Quant Time: Sep 20 01:32:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	148668	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	252923	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	214621	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	80416	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	116322	51.890	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.780%			
35) Dibromofluoromethane	8.165	113	80339	49.617	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.240%			
50) Toluene-d8	10.565	98	311024	55.564	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 111.120%			
62) 4-Bromofluorobenzene	12.847	95	106320	46.037	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 92.080%			
Target Compounds						
16) Acetone	4.424	43	17842	19.854	ug/l	91
20) Methylene Chloride	5.277	84	5760	3.132	ug/l	88
25) 2-Butanone	7.489	43	9347	7.914	ug/l #	90
43) Isopropyl Acetate	8.688	43	136382	34.021	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091924\
Data File : VN084036.D
Acq On : 19 Sep 2024 16:03
Operator : JC\MD
Sample : P4085-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-2

Quant Time: Sep 20 01:32:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

