

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091924\
 Data File : VN084037.D
 Acq On : 19 Sep 2024 16:28
 Operator : JC\MD
 Sample : P4003-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW178I-20240911

Quant Time: Sep 20 01:33:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	159427	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	275884	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	231845	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	97528	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	122896	51.123	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	102.240%	
35) Dibromofluoromethane	8.165	113	86477	48.963	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	97.920%	
50) Toluene-d8	10.565	98	325029	53.233	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	106.460%	
62) 4-Bromofluorobenzene	12.847	95	118268	46.949	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	93.900%	
Target Compounds						
44) Trichloroethene	9.353	130	1330	0.718	ug/l	Qvalue 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091924\
Data File : VN084037.D
Acq On : 19 Sep 2024 16:28
Operator : JC\MD
Sample : P4003-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW178I-20240911

Quant Time: Sep 20 01:33:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

