

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091924\
 Data File : VN084043.D
 Acq On : 19 Sep 2024 18:53
 Operator : JC\MD
 Sample : P4003-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE120D1-20240912

Quant Time: Sep 20 01:35:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	149356	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	256415	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	217092	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	87940	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	117367	52.115	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 104.220%		
35) Dibromofluoromethane	8.165	113	82554	50.291	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.580%		
50) Toluene-d8	10.565	98	308261	54.320	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 108.640%		
62) 4-Bromofluorobenzene	12.847	95	108647	46.404	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 92.800%		
Target Compounds						
44) Trichloroethene	9.353	130	25842	15.006	ug/l	Qvalue 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091924\
Data File : VN084043.D
Acq On : 19 Sep 2024 18:53
Operator : JC\MD
Sample : P4003-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE120D1-20240912

Quant Time: Sep 20 01:35:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

